



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118391

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	01/26/2022	11:25
2	N1253-01		1	12	100.00	>212.0	>212.0	01/26/2022	11:45
3	N1253-01DUP		1	12	100.00	>212.0	>212.0	01/26/2022	12:10
4	N1257-03		1	15	100.00	>212.0	>212.0	01/26/2022	12:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB118391

WorkList Name : FP12522 WorkList ID : 156307 Department : Wet-Chemistry Date : 01-25-2022 17:24:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/07/2022	Water	N1253-01	C Flash Point	Cool 4 deg C	PSEG03	D12	NWB-1732	01/24/2022	1010B
01/28/2022	Water	N1257-03	B Flash Point	Cool 4 deg C	PSEG03	L11	CPIT-COMP	01/25/2022	1010B

Date/Time 01/26/22 11:00
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 01/26/22 12:40
 Raw Sample Received by: RS C (Est-1066)
 Raw Sample Relinquished by: RS (WC)