



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118394

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/27/2022	10:15
2	CAL2	1	Water	NA	NA	20.2	7.00	01/27/2022	10:16
3	CAL3	1	Water	NA	NA	20.3	10.03	01/27/2022	10:17
4	ICV	1	Water	NA	NA	20.4	6.99	01/27/2022	10:19
5	CCV1	1	Water	NA	NA	20.6	2.01	01/27/2022	10:20
6	N1287-04	1	Water	NA	NA	20.2	7.60	01/27/2022	10:22
7	N1287-04DUP	1	Water	NA	NA	20.3	7.61	01/27/2022	10:23
8	CCV2	1	Water	NA	NA	20.2	12.03	01/27/2022	10:25

118394

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N1287 w WorkList ID : 156340 Department : Wet-Chemistry Date : 01-27-2022 10:02:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/25/2022	Water	N1287-04	Corrosivity	Cool 4 deg C	DVIR01	D11	WASTE-CHARACTER-2	01/25/2022	9040C

Date/Time 01.27.22 10:10
Raw Sample Received by: J.D. (WC)
Raw Sample Relinquished by: J.D. (WC)

Date/Time 01.27.22 11:20
Raw Sample Received by: J.D. (WC)
Raw Sample Relinquished by: J.D. (WC)