



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: jigneshDate: 01/28/2022OVEN IN TEMP1: 103 °COVEN OUT TEMP1: 104 °CDATE & TIME IN1: 01/27/2022 16:00 DATE & TIME OUT1: 01/28/2022 08:00OVEN IN TEMP2: 104 °COVEN OUT TEMP2: 104 °CDATE & TIME IN2: 01/28/2022 08:30 DATE & TIME OUT2: 01/28/2022 10:00Run Number: LB118400BalanceID: WC SC-6OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB118400BL	LB118400BL	96.3718	96.3718	100	96.3719	96.3719	96.3719	0.0001	1
2	N1301-01	55-ST-34	89.1483	89.1483	100	89.2656	89.2656	89.2656	0.1173	1173
3	N1301-01DUP	55-ST-34DUP	85.4416	85.4416	100	85.5635	85.5635	85.5635	0.1219	1219

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

LB 118400

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS W N1301 WorkList ID : 156354 Department : Wet-Chemistry Date : 01-27-2022 11:27:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/03/2022	Water	N1301-01	TS ~✓	Cool 4 deg C	DIST01	K11	55-ST-34	01/27/2022	SM2540 B

Date/Time 01.27.22 15:00

Raw Sample Received by: JB (607)

Raw Sample Relinquished by: CL 302

Date/Time 01.27.22

Raw Sample Received by: CL 302

Raw Sample Relinquished by: JB (607)