



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 1/28/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:30
In Date: 01/27/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:28
Out Date: 01/28/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB118405

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1290-01	BGYS2	1	1.12	8.63	9.75	7.72	76.5	
N1290-02	BGYS3	2	1.12	8.68	9.8	7.87	77.8	
N1290-03	BGYS4	3	1.12	8.86	9.98	8.88	87.6	
N1290-04	BGYS5	4	1.12	8.73	9.85	7.85	77.1	
N1290-05	BGYS6	5	1.12	8.68	9.8	8.17	81.2	
N1290-06	BGYS7	6	1.12	8.82	9.94	9.14	90.9	
N1290-07	BGYS8	7	1.12	8.73	9.85	9.1	91.4	
N1290-08	BGYS9	8	1.12	8.72	9.84	8.98	90.1	
N1290-09	BGYT0	9	1.12	8.84	9.96	9.14	90.7	
N1290-10	BGYT1	10	1.12	8.78	9.9	7.6	73.8	
N1290-11	BGYT2	11	1.12	8.86	9.98	8.88	87.6	
N1290-12	BGYT3	12	1.12	8.83	9.95	8.79	86.9	
N1290-13	BGYT4	13	1.12	8.85	9.97	9.06	89.7	
N1290-14	BGYT5	14	1.12	8.77	9.89	9.16	91.7	
N1290-15	BGYT6	15	1.12	8.66	9.78	8.99	90.9	
N1290-16	BGYT7	16	1.12	8.77	9.89	9.03	90.2	
N1290-17	BGYT8	17	1.12	8.63	9.75	7.22	70.7	
N1290-18	BGYT9	18	1.12	8.84	9.96	9.02	89.4	
N1290-19	BGYW0	19	1.12	8.86	9.98	9.2	91.2	
N1290-20	BGYW0D	20	1.12	8.86	9.98	9.2	91.2	
N1290-21	BGYW0S	21	1.12	8.86	9.98	9.2	91.2	
N1290-22	BGYW1	22	1.12	8.70	9.82	9.05	91.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$