


LB112409

7470A

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 2.6877e-004

C= -4.8770e-002

Rho= 0.9998925

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	ppb
0.0	0.000	-0.041	-0.041	29	0.000	29					-17
0.2	0.200	0.166	-0.034	799	0.0 %	799					4
2.5	2.500	2.602	0.102	9862	0.0 %	9862					1
5.0	5.000	5.029	0.029	18892	0.0 %	18892					1
7.5	7.500	7.441	-0.059	27866	0.0 %	27866					1
10.0	10.000	10.003	0.003	37400	0.0 %	37400					0

LB118407 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Jan 2022 13:31:19

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	5.0	29	-	0.00007470A	PPB	27 Jan 2022 13:43:03	S	Std
0.2 - 1	50.2	799	-	0.20007470A	PPB	27 Jan 2022 13:45:20	S	Std
2.5 - 1	52.5	9862	-	2.50007470A	PPB	27 Jan 2022 13:47:37	S	Std
5.0 - 1	55	18892	-	5.00007470A	PPB	27 Jan 2022 13:49:54	S	Std
7.5 - 1	57.5	27866	-	7.50007470A	PPB	27 Jan 2022 13:52:10	S	Std
10.0 - 1	51.0	37400	-	10.00007470A	PPB	27 Jan 2022 13:56:57	S	Std
ICV62 - 1	ICV62	14913	3.9594	-7470A	PPB	27 Jan 2022 14:04:17	U	SMPL
ICB62 - 1	ICB62	-61	-0.0652	-7470A	PPB	27 Jan 2022 14:06:32	U	SMPL
CCV76 - 1	CCV76	18574	4.9434	-7470A	PPB	27 Jan 2022 14:08:47	U	SMPL
CCB76 - 1	CCB76	-31	-0.0571	-7470A	PPB	27 Jan 2022 14:11:02	U	SMPL
CRA - 1	CRA	766	0.1571	-7470A	PPB	27 Jan 2022 14:13:18	U	SMPL
HighStd - 1	HighStd	36730	9.8233	-7470A	PPB	27 Jan 2022 14:15:32	U	SMPL
ChkStd - 1	ChkStd	24096	6.4276	-7470A	PPB	27 Jan 2022 14:17:48	U	SMPL
PB142319BL - 1	PBW	-117	-0.0802	-7470A	PPB	27 Jan 2022 14:20:05	U	SMPL
PB142319BS - 1	LCSW	13398	3.5523	-7470A	PPB	27 Jan 2022 14:24:54	U	SMPL
N1287-04 - 1	WASTE-CHARACTER-2	-10	-0.0515	-7470A	PPB	27 Jan 2022 14:27:09	U	SMPL
N1287-04DUP - 1	WASTE-CHARACTER-2DUP	7	-0.0469	-7470A	PPB	27 Jan 2022 14:29:26	U	SMPL
N1287-04MS - 1	WASTE-CHARACTER-2MS	14642	3.8866	-7470A	PPB	27 Jan 2022 14:34:23	U	SMPL
N1287-04MSD - 1	WASTE-CHARACTER-2MSD	16445	4.3712	-7470A	PPB	27 Jan 2022 14:36:39	U	SMPL
CCV77 - 1	CCV77	17925	4.7690	-7470A	PPB	27 Jan 2022 14:38:55	U	SMPL
CCB77 - 1	CCB77	-35	-0.0582	-7470A	PPB	27 Jan 2022 14:41:10	U	SMPL
N1287-04LX5 - 1		0	-0.0488	-7470A	PPB	27 Jan 2022 14:48:04	U	SMPL
N1287-04A - 1		15632	4.1527	-7470A	PPB	27 Jan 2022 14:52:50	U	SMPL
PB142313TB - 1	PB142313TB	-35	-0.0582	-7470A	PPB	27 Jan 2022 14:55:04	U	SMPL
CCV78 - 1	CCV78	18011	4.7921	-7470A	PPB	27 Jan 2022 14:57:20	U	SMPL
CCB78 - 1	CCB78	-24	-0.0552	-7470A	PPB	27 Jan 2022 14:59:36	U	SMPL