

LB118418


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

55979



A= 0.0000e+000

B= 1.7965e-004

C= 1.5617e-002

Rho= 0.9997110

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.0	0.000	0.023	0.023	39	0.000	39					
0.2	0.200	0.212	0.012	1095	0.0 %	1095					
2.5	2.500	2.437	-0.063	13480	0.0 %	13480					
5.0	5.000	5.113	0.113	28372	0.0 %	28372					
7.5	7.500	7.343	-0.157	40786	0.0 %	40786					
10.0	10.000	10.072	0.072	55979	0.0 %	55979					

LB118418 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 28 Jan 2022 11:06:01					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	39	-	0.0000	7471B	PPB	28 Jan 2022 12:16:12	S	Std
0.2 - 1	50.2	1095	-	0.2000	7471B	PPB	28 Jan 2022 12:18:29	S	Std
2.5 - 1	52.5	13480	-	2.5000	7471B	PPB	28 Jan 2022 12:20:46	S	Std
5.0 - 1	55	28372	-	5.0000	7471B	PPB	28 Jan 2022 12:26:58	S	Std
7.5 - 1	57.5	40786	-	7.5000	7471B	PPB	28 Jan 2022 12:29:14	S	Std
10.0 - 1	510	55979	-	10.0000	7471B	PPB	28 Jan 2022 12:34:04	S	Std
ICV64 - 1	ICV64	21053	3.7978	-	7471B	PPB	28 Jan 2022 12:39:19	U	SMPL
ICB64 - 1	ICB64	-40	0.0084	-	7471B	PPB	28 Jan 2022 12:41:33	U	SMPL
CCV82 - 1	CCV82	28534	5.1418	-	7471B	PPB	28 Jan 2022 12:43:49	U	SMPL
CCB82 - 1	CCB82	-73	0.0025	-	7471B	PPB	28 Jan 2022 12:46:04	U	SMPL
CRA - 1	CRA	1007	0.1965	-	7471B	PPB	28 Jan 2022 12:48:20	U	SMPL
HighStd - 1	HighStd	56083	10.0909	-	7471B	PPB	28 Jan 2022 12:50:35	U	SMPL
ChkStd - 1	ChkStd	37280	6.7130	-	7471B	PPB	28 Jan 2022 12:52:50	U	SMPL
PB142345BL - 1	PBS	-61	0.0047	-	7471B	PPB	28 Jan 2022 12:55:08	U	SMPL
PB142345BS - 1	LCSS	18741	3.3824	-	7471B	PPB	28 Jan 2022 12:59:56	U	SMPL
N1298-01 - 1	SPAT-SOIL-CUT	1081	0.2098	-	7471B	PPB	28 Jan 2022 13:02:12	U	SMPL
N1298-01DUP - 1	SPAT-SOIL-CUTDUP	1081	0.2098	-	7471B	PPB	28 Jan 2022 13:04:29	U	SMPL
N1298-01MS - 1	SPAT-SOIL-CUTMS	20053	3.6181	-	7471B	PPB	28 Jan 2022 13:06:45	U	SMPL
N1298-01MSD - 1	SPAT-SOIL-CUTMSD	21271	3.8370	-	7471B	PPB	28 Jan 2022 13:09:03	U	SMPL
N1309-11 - 1	STATE-ST-SP	494	0.1044	-	7471B	PPB	28 Jan 2022 13:11:18	U	SMPL
CCV83 - 1	CCV83	28838	5.1964	-	7471B	PPB	28 Jan 2022 13:13:34	U	SMPL
CCB83 - 1	CCB83	-43	0.0079	-	7471B	PPB	28 Jan 2022 13:15:50	U	SMPL
N1310-01 - 1	SOIL-PILE	3185	0.5878	-	7471B	PPB	28 Jan 2022 13:18:05	U	SMPL
N1312-01 - 1	OR-03-012722	1815	0.3417	-	7471B	PPB	28 Jan 2022 13:20:21	U	SMPL
N1314-01 - 1	CL-01-012722	5638	1.0285	-	7471B	PPB	28 Jan 2022 13:22:36	U	SMPL
N1315-01 - 1	OK-02-012722	2061	0.3859	-	7471B	PPB	28 Jan 2022 13:24:52	U	SMPL
N1316-01 - 1	TR-05-012722	2280	0.4252	-	7471B	PPB	28 Jan 2022 13:27:08	U	SMPL
N1319-01 - 1	MH-1S	2092	0.3914	-	7471B	PPB	28 Jan 2022 13:29:24	U	SMPL
N1319-04 - 1	MH-1D	826	0.1640	-	7471B	PPB	28 Jan 2022 13:31:40	U	SMPL
N1319-07 - 1	MH-2S	810	0.1611	-	7471B	PPB	28 Jan 2022 13:33:58	U	SMPL
N1319-10 - 1	MH-2D	713	0.1437	-	7471B	PPB	28 Jan 2022 13:36:15	U	SMPL
CCV84 - 1	CCV84	28730	5.1770	-	7471B	PPB	28 Jan 2022 13:38:30	U	SMPL
CCB84 - 1	CCB84	-29	0.0104	-	7471B	PPB	28 Jan 2022 13:40:46	U	SMPL
N1298-01LX5 - 1		189	0.0496	-	7471B	PPB	28 Jan 2022 13:43:01	U	SMPL
N1298-01A - 1		20712	3.7365	-	7471B	PPB	28 Jan 2022 13:45:17	U	SMPL
CCV85 - 1	CCV85	28880	5.2039	-	7471B	PPB	28 Jan 2022 13:47:32	U	SMPL
CCB85 - 1	CCB85	-35	0.0093	-	7471B	PPB	28 Jan 2022 13:49:48	U	SMPL