



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118420

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/28/2022	10:15
2	CAL2	1	Water	NA	NA	20.3	7.00	01/28/2022	10:16
3	CAL3	1	Water	NA	NA	20.3	10.03	01/28/2022	10:17
4	ICV	1	Water	NA	NA	20.4	6.99	01/28/2022	10:19
5	CCV1	1	Water	NA	NA	20.3	2.01	01/28/2022	10:20
6	N1284-01	1	Solid	20.02	20	20.4	8.36	01/28/2022	10:25
7	N1309-11	1	Solid	20.03	20	20.3	7.77	01/28/2022	10:26
8	N1313-01	1	Solid	20.02	20	20.6	8.70	01/28/2022	10:42
9	N1313-01DUP	1	Solid	20.03	20	20.7	8.71	01/28/2022	10:43
10	CCV2	1	Water	NA	NA	20.4	12.03	01/28/2022	10:44

VB 118420

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n13213 WorkList ID : 156391 Department : Wet-Chemistry Date : 01-28-2022 09:44:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/25/2022	Solid	N1284-01	pH	Cool 4 deg C	DIST01	H11	S-1	01/25/2022	9045D
01/27/2022	Solid	N1309-11	pH	Cool 4 deg C	NEWP02	K11	STATE-ST-SP	01/27/2022	9045D
01/27/2022	Solid	N1313-01	pH	Cool 4 deg C	WALS01	K11	17B-MISC-A	01/27/2022	9045D

Date/Time 01-28-22 10:10
Raw Sample Received by: JD (QC)
Raw Sample Relinquished by: CP SN

Date/Time 01-28-22 12:30
Raw Sample Received by: CP SN
Raw Sample Relinquished by: JD (QC)