



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 1/31/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:05
In Date: 01/28/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:05
Out Date: 01/29/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB118423

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1292-01	BGYY1	1	1.12	8.83	9.95	8.74	86.3	
N1292-02	BGYY2	2	1.12	8.59	9.71	5.5	51.0	
N1292-03	BGYY3	3	1.12	8.78	9.9	8.21	80.8	
N1292-04	BGYY4	4	1.12	8.76	9.88	8.4	83.1	
N1292-05	BGYY5	5	1.12	8.86	9.98	4.02	32.7	
N1292-06	BGYY6	6	1.12	8.70	9.82	8.06	79.8	
N1292-07	BGYY7	7	1.12	8.83	9.95	8.74	86.3	
N1292-08	BGYY9	8	1.12	8.81	9.93	8.98	89.2	
N1292-09	BGYY9D	9	1.12	8.81	9.93	8.98	89.2	
N1292-10	BGYY9S	10	1.12	8.81	9.93	8.98	89.2	
N1292-11	BGYZ0	11	1.12	8.73	9.85	8.37	83.0	
N1292-12	BGYZ1	12	1.12	8.86	9.98	8.59	84.3	
N1292-13	BGZ04	13	1.12	8.78	9.9	9.08	90.7	
N1292-14	BGZ08	14	1.12	8.73	9.85	8.07	79.6	
N1292-15	BGZ09	15	1.12	8.66	9.78	6.39	60.9	
N1292-16	BGZ10	16	1.12	8.84	9.96	9.04	89.6	
N1292-17	BGZ18	17	1.12	8.85	9.97	9.04	89.5	
N1292-18	BGZ20	18	1.12	8.78	9.9	9.16	91.6	
N1292-19	BGZ26	19	1.12	8.76	9.88	8.27	81.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$