


LB118429

245.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.8520e-004

C= -7.9289e-003

Rho= 0.9999714

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.0	0.000	0.009	0.009	59	0.000	59					
0.05	0.050	0.047	-0.003	194	0.0 %	194	0				
0.2	0.200	0.203	0.003	740	0.0 %	740					
2.5	2.500	2.500	0.000	8795	0.0 %	8795					
5.0	5.000	5.012	0.012	17603	0.0 %	17603					
7.5	7.500	7.439	-0.061	26111	0.0 %	26111					
10.0	10.000	10.039	0.039	35228	0.0 %	35228					

0.0002

LB118429 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 28 Jan 2022 14:07:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	59	-	0.0000	245.1	PPB	28 Jan 2022 14:22:14	S	Std
0.05 - 1	50.05	194	-	0.0500	245.1	PPB	28 Jan 2022 14:27:03	S	Std
0.2 - 1	502	740	-	0.2000	245.1	PPB	28 Jan 2022 14:29:20	S	Std
2.5 - 1	52.5	8795	-	2.5000	245.1	PPB	28 Jan 2022 14:31:37	S	Std
5.0 - 1	55	17603	-	5.0000	245.1	PPB	28 Jan 2022 14:33:53	S	Std
7.5 - 1	57.5	26111	-	7.5000	245.1	PPB	28 Jan 2022 14:36:09	S	Std
10.0 - 1	510	35228	-	10.0000	245.1	PPB	28 Jan 2022 14:38:25	S	Std
ICV65 - 1	ICV65	13698	3.8987	-245.1		PPB	28 Jan 2022 14:41:17	U	SMPL
ICB65 - 1	ICB65	-11	-0.0111	-245.1		PPB	28 Jan 2022 14:43:31	U	SMPL
CCV86 - 1	CCV86	17803	5.0694	-245.1		PPB	28 Jan 2022 14:45:46	U	SMPL
CCB86 - 1	CCB86	-12	-0.0114	-245.1		PPB	28 Jan 2022 14:48:02	U	SMPL
CRA - 1	CRA	178	0.0428	-245.1		PPB	28 Jan 2022 14:50:16	U	SMPL
HighStd - 1	HighStd	35529	10.1248	-245.1		PPB	28 Jan 2022 14:52:31	U	SMPL
ChkStd - 1	ChkStd	24706	7.0381	-245.1		PPB	28 Jan 2022 14:54:47	U	SMPL
PB142363BL - 1	PBW	-54	-0.0233	-245.1		PPB	28 Jan 2022 14:57:04	U	SMPL
PB142363BS - 1	LCSW	13769	3.9189	-245.1		PPB	28 Jan 2022 15:01:53	U	SMPL
N1202-01 - 1	MANHOLE	77	0.0140	-245.1		PPB	28 Jan 2022 15:04:09	U	SMPL
N1202-01DUP - 1	MANHOLEDUP	134	0.0303	-245.1		PPB	28 Jan 2022 15:06:25	U	SMPL
N1202-01MS - 1	MANHOLEMS	12450	3.5428	-245.1		PPB	28 Jan 2022 15:08:43	U	SMPL
N1202-01MSD - 1	MANHOLEMSD	12093	3.4410	-245.1		PPB	28 Jan 2022 15:11:00	U	SMPL
CCV87 - 1	CCV87	17859	5.0854	-245.1		PPB	28 Jan 2022 15:15:31	U	SMPL
CCB87 - 1	CCB87	-10	-0.0108	-245.1		PPB	28 Jan 2022 15:17:46	U	SMPL
N1205-05 - 1	PT-HG-WS	12525	3.5642	-245.1		PPB	28 Jan 2022 15:24:32	U	SMPL
N1301-01 - 1	55-ST-34	-21	-0.0139	-245.1		PPB	28 Jan 2022 15:26:48	U	SMPL
N1202-01LX5 - 1		65	0.0106	-245.1		PPB	28 Jan 2022 15:29:04	U	SMPL
N1202-01A - 1		13649	3.8847	-245.1		PPB	28 Jan 2022 15:31:20	U	SMPL
CCV88 - 1	CCV88	17837	5.0791	-245.1		PPB	28 Jan 2022 15:33:37	U	SMPL
CCB88 - 1	CCB88	11	-0.0048	-245.1		PPB	28 Jan 2022 15:35:54	U	SMPL