



Analytical Summary Report

Analysis Method: 9045D
Parameter: Corrosivity
Run Number: LB118438
BalanceID: WC SC-4

Analyst By : jignesh
Supervisor Review By : apatel
Slope : 98.7
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/28/2022	17:15
2	CAL2	1	Water	NA	NA	20.3	7.00	01/28/2022	17:16
3	CAL3	1	Water	NA	NA	20.3	10.03	01/28/2022	17:17
4	ICV	1	Water	NA	NA	20.4	7.02	01/28/2022	17:19
5	CCV1	1	Water	NA	NA	20.3	2.01	01/28/2022	17:20
6	N1325-02	1	Solid	20.03	20	20.4	6.20	01/28/2022	17:22
7	N1325-02DUP	1	Solid	20.04	20	20.5	6.21	01/28/2022	17:23
8	N1325-05	1	Solid	20.03	20	20.3	6.70	01/28/2022	17:25
9	N1325-08	1	Solid	20.04	20	20.4	7.01	01/28/2022	17:26
10	N1325-11	1	Solid	20.03	20	20.3	7.72	01/28/2022	17:29
11	N1325-14	1	Solid	20.05	20	20.5	7.54	01/28/2022	17:30
12	N1325-17	1	Solid	20.04	20	20.2	7.45	01/28/2022	17:32
13	N1325-20	1	Solid	20.01	20	20.3	7.30	01/28/2022	17:35
14	N1335-02	1	Solid	20.03	20	20.5	9.30	01/28/2022	17:39
15	N1335-05	1	Solid	20.02	20	20.5	8.81	01/28/2022	17:42
16	CCV2	1	Water	NA	NA	20.6	12.03	01/28/2022	17:44
17	N1335-08	1	Solid	20.02	20	20.4	8.54	01/28/2022	17:46
18	N1335-11	1	Solid	20.03	20	20.2	7.81	01/28/2022	17:48
19	CCV3	1	Water	NA	NA	20.5	2.02	01/28/2022	17:49

WB 118438

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N1325 WorkList ID : 156410 Department : Wet-Chemistry Date : 01-28-2022 14:21:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/28/2022	Solid	N1325-02	Corrosivity	Cool 4 deg C	PSEG03	K11	TES-1	01/28/2022	9045D
01/28/2022	Solid	N1325-05	Corrosivity	Cool 4 deg C	PSEG03	K11	TES-2	01/28/2022	9045D
01/28/2022	Solid	N1325-08	Corrosivity	Cool 4 deg C	PSEG03	K11	TES-3	01/28/2022	9045D
01/28/2022	Solid	N1325-11	Corrosivity	Cool 4 deg C	PSEG03	K11	TES-4	01/28/2022	9045D
01/28/2022	Solid	N1325-14	Corrosivity	Cool 4 deg C	PSEG03	K11	TEX-1	01/28/2022	9045D
01/28/2022	Solid	N1325-17	Corrosivity	Cool 4 deg C	PSEG03	K11	TEX-2	01/28/2022	9045D
01/28/2022	Solid	N1325-20	Corrosivity	Cool 4 deg C	PSEG03	K11	TEX-3	01/28/2022	9045D
01/28/2022	Solid	N1335-02	Corrosivity	Cool 4 deg C	PSEG03	K11	MH-3S	01/28/2022	9045D
01/28/2022	Solid	N1335-05	Corrosivity	Cool 4 deg C	PSEG03	K11	MH-3D	01/28/2022	9045D
01/28/2022	Solid	N1335-08	Corrosivity	Cool 4 deg C	PSEG03	K11	MH-4S	01/28/2022	9045D
01/28/2022	Solid	N1335-11	Corrosivity	Cool 4 deg C	PSEG03	K11	MH-4D	01/28/2022	9045D

Date/Time 01-28-22 15:30
 Raw Sample Received by: JO (GJC)
 Raw Sample Relinquished by: CP SM

Date/Time 01-28-22 18:10
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JO (GJC)