



# PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 2/17/2022

OVENTEMP IN Celsius(°C): 107  
Time IN: 14:20  
In Date: 01/31/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:25  
Out Date: 02/01/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB118450

| Lab ID   | Client SampleID | Dish # | Dish Wt (g) (A) | Sample Wt (g) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments    |
|----------|-----------------|--------|-----------------|---------------|--------------------------|----------------------------|---------|-------------|
| N1326-01 | GBCP7           | 1      | 1.13            | 8.43          | 9.56                     | 9.55                       | 99.9    |             |
| N1326-02 | GBCP8           | 2      | 1.11            | 8.76          | 9.87                     | 9.8                        | 99.2    |             |
| N1326-03 | GBCP9           | 3      | 1.15            | 8.83          | 9.98                     | 9.95                       | 99.7    |             |
| N1326-04 | GBCQ0           | 4      | 1.16            | 8.42          | 9.58                     | 8.53                       | 87.5    |             |
| N1326-05 | GBCQ1           | 5      | 1.18            | 8.64          | 9.82                     | 9.66                       | 98.1    |             |
| N1326-06 | GBCQ1D          | 6      | 1.18            | 8.64          | 9.82                     | 9.66                       | 98.1    |             |
| N1326-07 | GBCQ1S          | 7      | 1.18            | 8.64          | 9.82                     | 9.66                       | 98.1    |             |
| N1326-08 | GBCQ2           | 8      | 1.13            | 8.62          | 9.75                     | 9.7                        | 99.4    |             |
| N1326-09 | GBCQ3           | 9      | 1.16            | 8.77          | 9.93                     | 8.81                       | 87.2    |             |
| N1326-10 | GBCQ4           | 10     | 1.12            | 8.78          | 9.9                      | 9.55                       | 96.0    |             |
| N1326-11 | GBCQ5           | 11     | 1.1             | 8.77          | 9.87                     | 9.47                       | 95.4    |             |
| N1326-12 | GBCQ6           | 12     | 1.19            | 8.61          | 9.8                      | 9.55                       | 97.1    |             |
| N1326-13 | GBCQ7           | 13     | 1.13            | 8.50          | 9.63                     | 9.35                       | 96.7    |             |
| N1326-14 | GBCQ8           | 14     | 1.18            | 8.51          | 9.69                     | 9.09                       | 92.9    |             |
| N1326-15 | GBCQ9           | 15     | 1.18            | 8.55          | 9.73                     | 9.22                       | 94.0    |             |
| N1326-16 | GBCR0           | 16     | 1.19            | 8.67          | 9.86                     | 8.82                       | 88.0    |             |
| N1326-17 | GBCR1           | 17     | 1.15            | 8.60          | 9.75                     | 8.1                        | 80.8    |             |
| N1326-18 | GBCR2           | 18     | 1.12            | 8.63          | 9.75                     | 8.82                       | 89.2    |             |
| N1326-19 | GBCR3           | 19     | 1.19            | 8.63          | 9.82                     | 9.69                       | 98.5    |             |
| N1326-20 | GBCR4           | 20     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | P.E. SAMPLE |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$