



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118460

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2743     |

| Seq | LabID       | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 8                 | 27.00      | 80.6      | 80.3            | 01/31/2022 | 16:50     |
| 2   | N1342-29    |               | 1  | 12                | 100.00     | >212.0    | >212.0          | 01/31/2022 | 17:10     |
| 3   | N1342-29DUP |               | 1  | 12                | 100.00     | >212.0    | >212.0          | 01/31/2022 | 17:35     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

# WORKLIST(Hardcopy Internal Chain)

LB 118460

WorkList Name : FP13122      WorkList ID : 156453      Department : Wet-Chemistry      Date : 01-31-2022 16:28:04

| Due Date   | Matrix | Sample   | Test          | Preservative | Customer | Raw Sample Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|---------------|--------------|----------|-----------------------------|-----------------|--------------|--------|
| 02/14/2022 | Solid  | N1342-29 | ✓ Flash Point | Cool 4 deg C | PSEG03   | K11                         | 31459           | 01/31/2022   | 1010A  |

Date/Time 01/31/22 16:30  
 Raw Sample Received by: KS (WC)  
 Raw Sample Relinquished by: AP SM

Date/Time 01/31/22 17:40  
 Raw Sample Received by: AP SM  
 Raw Sample Relinquished by: KS (WC)