


LB118464

7471B

INSTRUMENT ID: CVL



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/o D
0.0	0.000	-0.088	-0.088	249	0.000	249					1
0.2	0.200	0.221	0.021	1996	0.0 %	1996					11
2.5	2.500	2.548	0.048	15143	0.0 %	15143					2
5.0	5.000	5.102	0.102	29577	0.0 %	29577					2
7.5	7.500	7.435	-0.065	42755	0.0 %	42755					1
10.0	10.000	9.983	-0.017	57153	0.0 %	57153					0

LB118464 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Feb 2022 09:43:38

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	249	-	0.0000	7471B	PPB	01 Feb 2022 10:28:11	S	Std
0.2 - 1	50.2	1996	-	0.2000	7471B	PPB	01 Feb 2022 10:30:28	S	Std
2.5 - 1	52.5	15143	-	2.5000	7471B	PPB	01 Feb 2022 10:38:27	S	Std
5.0 - 1	85	29577	-	5.0000	7471B	PPB	01 Feb 2022 10:40:44	S	Std
7.5 - 1	57.5	42755	-	7.5000	7471B	PPB	01 Feb 2022 10:45:36	S	Std
10.0 - 1	51.0	57153	-	10.0000	7471B	PPB	01 Feb 2022 10:50:25	S	Std
ICV67 - 1	ICV67	23956	4.1075	-	7471B	PPB	01 Feb 2022 10:56:16	U	SMPL
ICB67 - 1	ICB67	317	-0.0764	-	7471B	PPB	01 Feb 2022 10:58:31	U	SMPL
CCV93 - 1	CCV93	29881	5.1561	-	7471B	PPB	01 Feb 2022 11:00:47	U	SMPL
CCB93 - 1	CCB93	310	-0.0776	-	7471B	PPB	01 Feb 2022 11:03:02	U	SMPL
CRA - 1	CRA	2035	0.2277	-	7471B	PPB	01 Feb 2022 11:05:18	U	SMPL
HighStd - 1	HighStd	58080	10.1470	-	7471B	PPB	01 Feb 2022 11:07:33	U	SMPL
ChkStd - 1	ChkStd	39513	6.8609	-	7471B	PPB	01 Feb 2022 11:09:48	U	SMPL
PB142399BL - 1	PBS	428	-0.0567	-	7471B	PPB	01 Feb 2022 11:14:40	U	SMPL
PB142399BS - 1	LCSS	20116	3.4278	-	7471B	PPB	01 Feb 2022 11:16:56	U	SMPL
N1303-01 - 1	CC-5	337770	59.6489	-	7471B	PPB	01 Feb 2022 11:19:12	U	SMPL
N1303-03 - 1	CC-6	266464	47.0286	-	7471B	PPB	01 Feb 2022 11:21:29	U	SMPL
N1303-05 - 1	CC-7	568836	HIGH	-	7471B	PPB	01 Feb 2022 11:23:56	X	
N1303-07 - 1	CC-8	304570	53.7729	-	7471B	PPB	01 Feb 2022 11:25:51	U	SMPL
N1303-09 - 1	CC-9	650453	HIGH	-	7471B	PPB	01 Feb 2022 11:28:14	X	
CCV94 - 1	CCV94	28911	4.9844	-	7471B	PPB	01 Feb 2022 11:30:07	U	SMPL
CCB94 - 1	CCB94	407	-0.0604	-	7471B	PPB	01 Feb 2022 11:35:05	U	SMPL
N1313-01 - 1	17B-MISC-A	7731	1.2358	-	7471B	PPB	01 Feb 2022 11:37:20	U	SMPL
N1325-01 - 1	TES-1	10674	1.7567	-	7471B	PPB	01 Feb 2022 11:44:22	U	SMPL
N1325-01DUP - 1	TES-1DUP	10710	1.7631	-	7471B	PPB	01 Feb 2022 11:46:38	U	SMPL
N1325-01MS - 1	TES-1MS	33211	5.7455	-	7471B	PPB	01 Feb 2022 11:51:24	U	SMPL
N1325-01MSD - 1	TES-1MSD	33518	5.7998	-	7471B	PPB	01 Feb 2022 11:53:40	U	SMPL
N1325-04 - 1	TES-2	23676	4.0579	-	7471B	PPB	01 Feb 2022 11:55:57	U	SMPL
N1325-07 - 1	TES-3	2260	0.2675	-	7471B	PPB	01 Feb 2022 11:58:15	U	SMPL
N1325-10 - 1	TES-4	345613	HIGH	-	7471B	PPB	01 Feb 2022 12:00:32	X	
N1325-13 - 1	TEX-1	1639	0.1576	-	7471B	PPB	01 Feb 2022 12:02:24	U	SMPL
N1325-16 - 1	TEX-2	9138	1.4848	-	7471B	PPB	01 Feb 2022 12:04:42	U	SMPL
CCV95 - 1	CCV95	29647	5.1147	-	7471B	PPB	01 Feb 2022 12:06:58	U	SMPL
CCB95 - 1	CCB95	481	-0.0473	-	7471B	PPB	01 Feb 2022 12:11:49	U	SMPL
N1325-19 - 1	TEX-3	21421	3.6588	-	7471B	PPB	01 Feb 2022 12:14:04	U	SMPL
N1335-01 - 1	MH-3S	947	0.0351	-	7471B	PPB	01 Feb 2022 12:16:19	U	SMPL
N1335-04 - 1	MH-3D	1248	0.0884	-	7471B	PPB	01 Feb 2022 12:18:35	U	SMPL
N1335-07 - 1	MH-4S	990	0.0427	-	7471B	PPB	01 Feb 2022 12:20:51	U	SMPL
N1335-10 - 1	MH-4D	759	0.0019	-	7471B	PPB	01 Feb 2022 12:23:06	U	SMPL
N1336-01 - 1	HD-01-012822	2759	0.3558	-	7471B	PPB	01 Feb 2022 12:25:22	U	SMPL
N1338-01 - 1	HP-1	1270	0.0923	-	7471B	PPB	01 Feb 2022 12:27:39	U	SMPL
N1325-01LX5 - 1		1855	0.1958	-	7471B	PPB	01 Feb 2022 12:29:56	U	SMPL
N1325-01A - 1		34866	6.0384	-	7471B	PPB	01 Feb 2022 12:32:13	U	SMPL
CCV96 - 1	CCV96	30501	5.2659	-	7471B	PPB	01 Feb 2022 12:34:29	U	SMPL
CCB96 - 1	CCB96	638	-0.0196	-	7471B	PPB	01 Feb 2022 12:36:46	U	SMPL
N1303-01DLX10 - 1		37919	6.5788	-	7471B	PPB	01 Feb 2022 12:39:02	U	SMPL
N1303-03DLX10 - 1		29053	5.0096	-	7471B	PPB	01 Feb 2022 12:41:18	U	SMPL
N1303-05DLX10 - 1		103570	18.1982	-	7471B	PPB	01 Feb 2022 12:43:34	U	SMPL
N1303-07DLX10 - 1		33095	5.7250	-	7471B	PPB	01 Feb 2022 12:45:51	U	SMPL
N1303-09DLX10 - 1		137219	24.1537	-	7471B	PPB	01 Feb 2022 12:48:09	U	SMPL
N1325-10DLX10 - 1		51083	8.9086	-	7471B	PPB	01 Feb 2022 12:50:25	U	SMPL
N1303-05DL2X50 - 1		20334	3.4664	-	7471B	PPB	01 Feb 2022 12:52:44	U	SMPL
N1303-09DL2X50 - 1		27153	4.6733	-	7471B	PPB	01 Feb 2022 12:55:01	U	SMPL
CCV97 - 1	CCV97	29984	5.1744	-	7471B	PPB	01 Feb 2022 12:57:20	U	SMPL
CCB97 - 1	CCB97	471	-0.0491	-	7471B	PPB	01 Feb 2022 12:59:38	U	SMPL