



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118465

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/01/2022	12:25
2	CAL2	1	Water	NA	NA	20.1	7.01	02/01/2022	12:26
3	CAL3	1	Water	NA	NA	20.1	10.03	02/01/2022	12:27
4	ICV	1	Water	NA	NA	20.2	6.99	02/01/2022	12:30
5	CCV1	1	Water	NA	NA	20.2	2.01	02/01/2022	12:31
6	N1343-01	1	Water	NA	NA	20.2	6.32	02/01/2022	12:33
7	N1351-01	1	Water	NA	NA	20.2	4.50	02/01/2022	12:35
8	N1351-01DUP	1	Water	NA	NA	20.3	4.51	02/01/2022	12:36
9	CCV2	1	Water	NA	NA	20.3	12.03	02/01/2022	12:37

W3718465

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1351 WorkList ID : 156474 Department : Wet-Chemistry Date : 02-01-2022 12:14:28

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
01/28/2022	Water	N1343-01	pH	Cool 4 deg C	PSEG03	K11	01/28/2022	9040C
02/14/2022	Water	N1351-01	pH	Cool 4 deg C	STER05	D11	02/25/2022	9040C

Date/Time 02-01-22 12:20
Raw Sample Received by: TP (WC)
Raw Sample Relinquished by: KS (WC)

Date/Time 02-01-22 13:30
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: TP (WC)