

Analysis Method: SM2510 BANALYST: IwonaParameter: ConductanceSUPERVISOR REVIEW BY: apatelRun Number: LB118470Meter ID: WC CONDUCTANCE METER-1Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP92563

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB118470BL		1	20.8	0.68	0.74	02/01/2022	13:00
2	LB118470BS	1413	1	21.1	1328.00	1434.88	02/01/2022	13:03
3	N1351-01		1	20.3	2.54	2.79	02/01/2022	13:06
4	N1351-01DUP		1	20.2	2.55	2.81	02/01/2022	13:10

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$
--

WORKLIST(Hardcopy Internal Chain)

LB 118470

WorkList Name : CONDDUCTANCE-0201; WorkList ID : 156473 Department : Wet-Chemistry Date : 02-01-2022 11:48:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/14/2022	Water	N1351-01	Conductance	Cool 4 deg C	STER05	D11	POOL-WATER	02/25/2022	SM2510 B

Date/Time 02/01/22 12:00
 Raw Sample Received by: 12 (WC)
 Raw Sample Relinquished by: 12 (WC)

Date/Time 02/01/22 1:30
 Raw Sample Received by: 12 (WC)
 Raw Sample Relinquished by: 12 (WC)