


LB 118474

7470A

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 2.5195e-004

C= -5.0372e-002

Rho= 0.9998937

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Yob
0.0	0.000	-0.005	-0.005	180	0.000	180					15
0.2	0.200	0.230	0.030	1113	0.0 %	1113					2
2.5	2.500	2.543	0.043	10294	0.0 %	10294					-2
5.0	5.000	4.917	-0.083	19715	0.0 %	19715					-1
7.5	7.500	7.446	-0.054	29754	0.0 %	29754					1
10.0	10.000	10.069	0.069	40164	0.0 %	40164					

LB118474 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Feb 2022 14:22:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	180	-	0.0000	7470A	PPB	01 Feb 2022 14:24:21	S	Std
0.2 - 1	50.2	1113	-	0.2000	7470A	PPB	01 Feb 2022 14:28:49	S	Std
2.5 - 1	52.5	10294	-	2.5000	7470A	PPB	01 Feb 2022 14:31:06	S	Std
5.0 - 1	55	19715	-	5.0000	7470A	PPB	01 Feb 2022 14:36:04	S	Std
7.5 - 1	57.5	29754	-	7.5000	7470A	PPB	01 Feb 2022 14:43:34	S	Std
10.0 - 1	510	40164	-	10.0000	7470A	PPB	01 Feb 2022 14:51:00	S	Std
ICV68 - 1	ICV68	16154	4.0196	-7470A	PPB	PPB	01 Feb 2022 14:56:08	U	SMPL
ICB68 - 1	ICB68	121	-0.0199	-7470A	PPB	PPB	01 Feb 2022 15:01:14	U	SMPL
CCV01 - 1	CCV01	19708	4.9150	-7470A	PPB	PPB	01 Feb 2022 15:03:29	U	SMPL
CCB01 - 1	CCB01	29	-0.0431	-7470A	PPB	PPB	01 Feb 2022 15:05:45	U	SMPL
CRA - 1	CRA	1119	0.2316	-7470A	PPB	PPB	01 Feb 2022 15:08:00	U	SMPL
HighStd - 1	HighStd	39913	10.0057	-7470A	PPB	PPB	01 Feb 2022 15:10:14	U	SMPL
ChkStd - 1	ChkStd	29479	7.3768	-7470A	PPB	PPB	01 Feb 2022 15:12:30	U	SMPL
PB142397BL - 1	PBW	147	-0.0133	-7470A	PPB	PPB	01 Feb 2022 15:14:46	U	SMPL
PB142397BS - 1	LCSW	16672	4.1501	-7470A	PPB	PPB	01 Feb 2022 15:17:02	U	SMPL
N1325-02 - 1	TES-1	68	-0.0332	-7470A	PPB	PPB	01 Feb 2022 15:19:18	U	SMPL
N1325-05 - 1	TES-2	93	-0.0269	-7470A	PPB	PPB	01 Feb 2022 15:21:34	U	SMPL
N1325-08 - 1	TES-3	95	-0.0264	-7470A	PPB	PPB	01 Feb 2022 15:23:51	U	SMPL
N1325-11 - 1	TES-4	108	-0.0232	-7470A	PPB	PPB	01 Feb 2022 15:26:08	U	SMPL
N1325-14 - 1	TEX-1	98	-0.0257	-7470A	PPB	PPB	01 Feb 2022 15:28:24	U	SMPL
CCV02 - 1	CCV02	19669	4.9052	-7470A	PPB	PPB	01 Feb 2022 15:30:39	U	SMPL
CCB02 - 1	CCB02	8	-0.0484	-7470A	PPB	PPB	01 Feb 2022 15:32:55	U	SMPL
N1325-17 - 1	TEX-2	88	-0.0282	-7470A	PPB	PPB	01 Feb 2022 15:35:10	U	SMPL
N1325-20 - 1	TEX-3	105	-0.0239	-7470A	PPB	PPB	01 Feb 2022 15:37:26	U	SMPL
N1335-02 - 1	MH-3S	109	-0.0229	-7470A	PPB	PPB	01 Feb 2022 15:39:41	U	SMPL
N1335-05 - 1	MH-3D	102	-0.0247	-7470A	PPB	PPB	01 Feb 2022 15:41:57	U	SMPL
N1335-08 - 1	MH-4S	123	-0.0194	-7470A	PPB	PPB	01 Feb 2022 15:44:13	U	SMPL
N1335-11 - 1	MH-4D	125	-0.0189	-7470A	PPB	PPB	01 Feb 2022 15:46:29	U	SMPL
N1335-11DUP - 1	MH-4DDUP	118	-0.0206	-7470A	PPB	PPB	01 Feb 2022 15:48:45	U	SMPL
N1335-11MS - 1	MH-4DMS	15408	3.8317	-7470A	PPB	PPB	01 Feb 2022 15:51:02	U	SMPL
N1335-11MSD - 1	MH-4DMSD	16036	3.9899	-7470A	PPB	PPB	01 Feb 2022 15:53:19	U	SMPL
PB142398BL - 1	PBW	25	-0.0441	-7470A	PPB	PPB	01 Feb 2022 15:55:35	U	SMPL
CCV03 - 1	CCV03	18473	4.6039	-7470A	PPB	PPB	01 Feb 2022 15:57:50	U	SMPL
CCB03 - 1	CCB03	-4	-0.0514	-7470A	PPB	PPB	01 Feb 2022 16:00:06	U	SMPL
PB142398BS - 1	LCSW	16137	4.0153	-7470A	PPB	PPB	01 Feb 2022 16:02:21	U	SMPL
N1264-01 - 1	PH1-BOT-8	223	0.0058	-7470A	PPB	PPB	01 Feb 2022 16:04:37	U	SMPL
N1264-02 - 1	PH1-BOT-20	308	0.0272	-7470A	PPB	PPB	01 Feb 2022 16:06:53	U	SMPL
N1264-02DUP - 1	PH1-BOT-20DUP	249	0.0124	-7470A	PPB	PPB	01 Feb 2022 16:09:09	U	SMPL
N1264-02MS - 1	PH1-BOT-20MS	15074	3.7475	-7470A	PPB	PPB	01 Feb 2022 16:11:24	U	SMPL
N1264-02MSD - 1	PH1-BOT-20MSD	16268	4.0483	-7470A	PPB	PPB	01 Feb 2022 16:13:40	U	SMPL
-PB142368TB - 1	-PB142368TB	79	-0.0305	-7470A	PPB	PPB	01 Feb 2022 16:15:57	U	SMPL
PB142398TB - 1	PB142370TB	170	-0.0075	-7470A	PPB	PPB	01 Feb 2022 16:18:14	U	SMPL
CCV04 - 1	CCV04	19738	4.9226	-7470A	PPB	PPB	01 Feb 2022 16:20:31	U	SMPL
CCB04 - 1	CCB04	42	-0.0398	-7470A	PPB	PPB	01 Feb 2022 16:22:47	U	SMPL
N1335-11LX5 - 1		23	-0.0446	-7470A	PPB	PPB	01 Feb 2022 16:25:02	U	SMPL
N1335-11A - 1		18298	4.5598	-7470A	PPB	PPB	01 Feb 2022 16:27:18	U	SMPL
N1264-02LX5 - 1		22	-0.0448	-7470A	PPB	PPB	01 Feb 2022 16:29:34	U	SMPL
N1264-02A - 1		16442	4.0922	-7470A	PPB	PPB	01 Feb 2022 16:31:49	U	SMPL
CCV05 - 1	CCV05	19747	4.9249	-7470A	PPB	PPB	01 Feb 2022 16:34:05	U	SMPL
CCB05 - 1	CCB05	11	-0.0476	-7470A	PPB	PPB	01 Feb 2022 16:36:20	U	SMPL