

LB118486


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

4057E



A= 0.0000e+000

B= 2.4827e-004

C= 1.3357e-002

Rho= 0.9994283

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.032	0.032	74	0.000	74					
0.2	0.200	0.198	-0.002	742	0.0 %	742					1
2.5	2.500	2.431	-0.069	9737	0.0 %	9737					3
5.0	5.000	5.175	0.175	20790	0.0 %	20790					4
7.5	7.500	7.278	-0.222	29260	0.0 %	29260					3
10.0	10.000	10.087	0.087	40576	0.0 %	40576					1

LB118486 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Feb 2022 08:59:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	74	-	0.0000	7470A	PPB	02 Feb 2022 10:38:15	S	Std
0.2 - 1	50.2	742	-	0.2000	7470A	PPB	02 Feb 2022 10:40:32	S	Std
2.5 - 1	52.5	9737	-	2.5000	7470A	PPB	02 Feb 2022 10:42:49	S	Std
5.0 - 1	55	20790	-	5.0000	7470A	PPB	02 Feb 2022 10:45:06	S	Std
7.5 - 1	57.5	29260	-	7.5000	7470A	PPB	02 Feb 2022 10:47:22	S	Std
10.0 - 1	51.0	40576	-	10.0000	7470A	PPB	02 Feb 2022 10:53:24	S	Std
ICV69 - 1	ICV69	15539	3.8712	-	7470A	PPB	02 Feb 2022 10:56:31	U	SMPL
ICB69 - 1	ICB69	5	0.0146	-	7470A	PPB	02 Feb 2022 10:58:46	U	SMPL
CCV06 - 1	CCV06	20982	5.2226	-	7470A	PPB	02 Feb 2022 11:01:01	U	SMPL
CCB06 - 1	CCB06	-9	0.0111	-	7470A	PPB	02 Feb 2022 11:03:17	U	SMPL
CRA - 1	CRA	722	0.1926	-	7470A	PPB	02 Feb 2022 11:05:32	U	SMPL
HighStd - 1	HighStd	40038	9.9536	-	7470A	PPB	02 Feb 2022 11:07:46	U	SMPL
ChkStd - 1	ChkStd	27560	6.8557	-	7470A	PPB	02 Feb 2022 11:10:02	U	SMPL
PB142417BL - 1	PBW	-33	0.0052	-	7470A	PPB	02 Feb 2022 11:12:18	U	SMPL
PB142417BS - 1	LCSW	14805	3.6890	-	7470A	PPB	02 Feb 2022 11:14:34	U	SMPL
N1342-01 - 1	31523	26	0.0198	-	7470A	PPB	02 Feb 2022 11:16:50	U	SMPL
N1342-01DUP - 1	31523DUP	63	0.0290	-	7470A	PPB	02 Feb 2022 11:19:06	U	SMPL
N1342-01MS - 1	31523MS	13655	3.4035	-	7470A	PPB	02 Feb 2022 11:21:23	U	SMPL
N1342-01MSD - 1	31523MSD	13845	3.4507	-	7470A	PPB	02 Feb 2022 11:23:40	U	SMPL
N1342-03 - 1	31524	46	0.0248	-	7470A	PPB	02 Feb 2022 11:25:55	U	SMPL
CCV07 - 1	CCV07	20959	5.2169	-	7470A	PPB	02 Feb 2022 11:28:10	U	SMPL
CCB07 - 1	CCB07	-4	0.0124	-	7470A	PPB	02 Feb 2022 11:30:26	U	SMPL
N1342-05 - 1	31525	64	0.0292	-	7470A	PPB	02 Feb 2022 11:32:41	U	SMPL
N1342-07 - 1	31526	114	0.0417	-	7470A	PPB	02 Feb 2022 11:34:56	U	SMPL
N1342-09 - 1	31527	67	0.0300	-	7470A	PPB	02 Feb 2022 11:37:11	U	SMPL
N1342-11 - 1	31528	113	0.0414	-	7470A	PPB	02 Feb 2022 11:39:27	U	SMPL
N1342-13 - 1	31534	406	0.1142	-	7470A	PPB	02 Feb 2022 11:41:42	U	SMPL
N1342-15 - 1	31535	89	0.0355	-	7470A	PPB	02 Feb 2022 11:43:59	U	SMPL
N1342-17 - 1	31536	86	0.0347	-	7470A	PPB	02 Feb 2022 11:46:15	U	SMPL
N1342-19 - 1	31537	107	0.0399	-	7470A	PPB	02 Feb 2022 11:48:32	U	SMPL
N1342-21 - 1	31538	81	0.0335	-	7470A	PPB	02 Feb 2022 11:50:49	U	SMPL
N1342-23 - 1	31539	208	0.0650	-	7470A	PPB	02 Feb 2022 11:53:04	U	SMPL
CCV08 - 1	CCV08	20718	5.1570	-	7470A	PPB	02 Feb 2022 11:55:20	U	SMPL
CCB08 - 1	CCB08	10	0.0158	-	7470A	PPB	02 Feb 2022 11:57:35	U	SMPL
N1342-25 - 1	31540	70	0.0307	-	7470A	PPB	02 Feb 2022 11:59:51	U	SMPL
N1342-27 - 1	40553	67	0.0300	-	7470A	PPB	02 Feb 2022 12:02:07	U	SMPL
N1342-29 - 1	31459	266	0.0794	-	7470A	PPB	02 Feb 2022 12:04:23	U	SMPL
PB142418BL - 1	PBW	27	0.0201	-	7470A	PPB	02 Feb 2022 12:06:38	U	SMPL
PB142418BS - 1	LCSW	13880	3.4594	-	7470A	PPB	02 Feb 2022 12:08:54	U	SMPL
N1343-01 - 1	1-CHEVALIER-AVE	1611	0.4133	-	7470A	PPB	02 Feb 2022 12:11:10	U	SMPL
N1347-01 - 1	RW8-MW015-20220127	38	0.0228	-	7470A	PPB	02 Feb 2022 12:13:27	U	SMPL
N1347-02 - 1	RW8-MW01D1-20220127	50	0.0258	-	7470A	PPB	02 Feb 2022 12:15:44	U	SMPL
N1347-02DUP - 1	RW8-MW01D1-20220127DUP	52	0.0263	-	7470A	PPB	02 Feb 2022 12:18:01	U	SMPL
N1347-03 - 1	N1347-02MS	14047	3.5008	-	7470A	PPB	02 Feb 2022 12:23:00	U	SMPL
CCV09 - 1	CCV09	20915	5.2060	-	7470A	PPB	02 Feb 2022 12:25:16	U	SMPL
CCB09 - 1	CCB09	10	0.0158	-	7470A	PPB	02 Feb 2022 12:27:31	U	SMPL
N1347-04 - 1	N1347-02MSD	14305	3.5649	-	7470A	PPB	02 Feb 2022 12:32:43	U	SMPL
N1347-05 - 1	RW8-MW01D3-20220128	30	0.0208	-	7470A	PPB	02 Feb 2022 12:34:59	U	SMPL
N1347-06 - 1	RW8-MW01D3-20220128-FIELD-FILTERED	54	0.0268	-	7470A	PPB	02 Feb 2022 12:37:15	U	SMPL
N1347-07 - 1	RW8-MW01D2-20220128	51	0.0260	-	7470A	PPB	02 Feb 2022 12:39:30	U	SMPL
N1347-08 - 1	RW8-MW01-DUP-20220127	63	0.0290	-	7470A	PPB	02 Feb 2022 12:41:46	U	SMPL
PB142402TB - 1	PB142402TB	71	0.0310	-	7470A	PPB	02 Feb 2022 12:44:02	U	SMPL
N1342-01LX5 - 1		23	0.0191	-	7470A	PPB	02 Feb 2022 12:46:19	U	SMPL
N1342-01A - 1		15570	3.8789	-	7470A	PPB	02 Feb 2022 12:48:36	U	SMPL
N1347-02LX5 - 1		-24	0.0074	-	7470A	PPB	02 Feb 2022 12:50:53	U	SMPL
N1347-02A - 1		15785	3.9323	-	7470A	PPB	02 Feb 2022 12:53:09	U	SMPL
CCV10 - 1	CCV10	20859	5.1921	-	7470A	PPB	02 Feb 2022 12:55:25	U	SMPL
CCB10 - 1	CCB10	9	0.0156	-	7470A	PPB	02 Feb 2022 12:57:40	U	SMPL