



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118487

Slope : 97.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/01/2022	16:20
2	CAL2	1	Water	NA	NA	20.2	7.00	02/01/2022	16:21
3	CAL3	1	Water	NA	NA	20.2	10.03	02/01/2022	16:22
4	ICV	1	Water	NA	NA	20.3	6.99	02/01/2022	16:25
5	CCV1	1	Water	NA	NA	20.4	2.01	02/01/2022	16:26
6	N1344-01	1	Water	NA	NA	20.1	7.48	02/01/2022	16:29
7	N1344-01DUP	1	Water	NA	NA	20.2	7.49	02/01/2022	16:30
8	N1354-03	1	Water	NA	NA	20.1	7.66	02/01/2022	16:33
9	CCV2	1	Water	NA	NA	20.3	12.03	02/01/2022	16:35

VB 118487

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1354 W WorkList ID : 156484 Department : Wet-Chemistry Date : 02-01-2022 16:08:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/01/2022	Water	N1344-01	pH	Cool 4 deg C	PSEG03	K11	33386	02/01/2022	9040C
02/01/2022	Water	N1354-03	pH	Cool 4 deg C	PSEG03	K11	33421	02/01/2022	9040C

Date/Time 02-01-22 16:15
Raw Sample Received by: JD (WC)
Raw Sample Relinquished by: CD gm

Date/Time 02-01-22 18:00
Raw Sample Received by: CD gm
Raw Sample Relinquished by: JD (WC)