



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118488

Slope : 97.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/01/2022	16:20
2	CAL2	1	Water	NA	NA	20.2	7.00	02/01/2022	16:21
3	CAL3	1	Water	NA	NA	20.2	10.03	02/01/2022	16:22
4	ICV	1	Water	NA	NA	20.3	6.99	02/01/2022	16:25
5	CCV1	1	Water	NA	NA	20.3	2.01	02/02/2022	17:25
6	N1354-01	1	Solid	20.03	20	20.2	8.52	02/01/2022	17:26
7	N1358-01	1	Solid	20.04	20	20.4	8.56	02/01/2022	17:29
8	N1358-01DUP	1	Solid	20.03	20	20.5	8.57	02/01/2022	17:30
9	CCV2	1	Water	NA	NA	20.4	12.02	02/01/2022	17:32

VB 118488

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1354 S WorkList ID : 156485 Department : Wet-Chemistry Date : 02-01-2022 16:09:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Customer Sample	Collect Date	Method
						Storage	Location			
02/01/2022	Solid	N1354-01	pH	Cool 4 deg C	PSEG03	K11	34030		02/01/2022	9045D
02/01/2022	Solid	N1358-01	pH	Cool 4 deg C	PSEG05	M11	BU-2-020122		02/01/2022	9045D

Date/Time 02-01-22 17:10
Raw Sample Received by: CD 16C1
Raw Sample Relinquished by: CD SM

Date/Time 02-01-22 18:25
Raw Sample Received by: CD SM
Raw Sample Relinquished by: CD 16C1