



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118490

Slope : 97.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/01/2022	16:20
2	CAL2	1	Water	NA	NA	20.2	7.00	02/01/2022	16:21
3	CAL3	1	Water	NA	NA	20.2	10.03	02/01/2022	16:22
4	ICV	1	Water	NA	NA	20.3	6.99	02/01/2022	16:25
5	CCV1	1	Water	NA	NA	20.4	2.01	02/01/2022	17:25
6	N1356-02	1	Solid	20.03	20	20.3	9.18	02/01/2022	17:30
7	N1356-02DUP	1	Solid	20.04	20	20.4	9.19	02/01/2022	17:31
8	N1356-04	1	Solid	20.02	20	20.3	9.25	02/01/2022	17:33
9	N1356-06	1	Solid	20.05	20	20.3	9.04	02/01/2022	17:35
10	N1357-02	1	Solid	20.02	20	20.1	6.40	02/01/2022	17:36
11	CCV2	1	Water	NA	NA	20.3	12.02	02/01/2022	17:39

AP 118489

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY N1357 S WorkList ID : 156481 Department : Wet-Chemistry Date : 02-01-2022 14:51:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/01/2022	Solid	N1356-02	Corrosivity	Cool 4 deg C	PSEG03	M11	F-PHC	02/01/2022	9045D
02/01/2022	Solid	N1356-04	Corrosivity	Cool 4 deg C	PSEG03	M11	M-PHC	02/01/2022	9045D
02/01/2022	Solid	N1356-06	Corrosivity	Cool 4 deg C	PSEG03	M11	B-PHC	02/01/2022	9045D
02/01/2022	Solid	N1357-02	Corrosivity	Cool 4 deg C	SEAC01	N11	SOIL-CUTTINGS	02/01/2022	9045D

Date/Time 02.01.22 17:10
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)

Date/Time 02.01.22 18:25
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)