



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118495

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	02/02/2022	12:20
2	N1343-01		1	10	100.00	>212.0	>212.0	02/02/2022	12:45
3	N1343-01DUP		1	10	100.00	>212.0	>212.0	02/02/2022	13:05
4	N1344-01		1	15	100.00	>212.0	>212.0	02/02/2022	13:40
5	N1354-03		1	12	100.00	>212.0	>212.0	02/02/2022	14:05
6	N1357-01		1	13	100.00	>212.0	>212.0	02/02/2022	14:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB118495

WorkList Name : FP2222 WorkList ID : 156504 Department : Wet-Chemistry Date : 02-02-2022 11:48:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/07/2022	Water	N1343-01	Flash Point	Cool 4 deg C	PSEG03	K11	1-CHEVALIER-AVE	01/28/2022	1010B
02/08/2022	Water	N1344-01	Flash Point	Cool 4 deg C	PSEG03	K11	33386	02/01/2022	1010B
02/08/2022	Water	N1354-03	Flash Point	Cool 4 deg C	PSEG03	K11	33421	02/01/2022	1010B
02/14/2022	Water	N1357-01	Flash Point	Cool 4 deg C	SEAC01	N11	GROUNDWATER	02/01/2022	1010B

Date/Time 02/02/22 11:55
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RJ CEAT-1062

Date/Time 02/02/22 14:45
 Raw Sample Received by: RJ CEAT-1062
 Raw Sample Relinquished by: KS (WC)