



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 2/3/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 16:50
In Date: 02/02/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 02/03/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB118496

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1361-01	01-A-01-B-01-C	1	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N1362-01	RO-342	2	1.14	8.76	9.9	5.72	52.3	GEL MATRIX
N1362-02	RO-342	3	1.18	8.55	9.73	6.79	65.6	GEL MATRIX
N1362-03	RO-341	4	1.19	8.41	9.6	6.53	63.5	GEL MATRIX
N1362-04	RO-341	5	1.19	8.64	9.83	6.23	58.3	GEL MATRIX
N1367-01	31246	8	1.00	1.00	2.00	2.00	100.0	OILY DEBRIS
N1373-01	OR-3-020222	9	1.15	8.58	9.73	8.96	91.0	
N1373-02	OR-3-020222-EPH-2	10	1.18	8.46	9.64	8.55	87.1	
N1378-01	ASTORIA-YARD-PIT-11-STO CKPILE	6	1.14	8.50	9.64	8.19	82.9	
N1378-01D	ASTORIA-YARD-PIT-11-STO CKPILEDUP	7	1.17	8.49	9.66	8.21	82.9	
N1382-01	POST-EX-BUILDING-1	11	1.18	8.66	9.84	8.93	89.5	
N1382-02	POST-EX-BUILDING-2	12	1.11	8.73	9.84	8.89	89.1	
N1382-03	POST-EX-BUILDING-3	13	1.15	8.81	9.96	9.09	90.1	
N1382-04	WASTE-BUILDING	14	1.19	8.61	9.8	8.92	89.8	
N1382-05	WASTE-BUILDING	15	1.19	8.61	9.8	8.92	89.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$