



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118500

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/02/2022	17:00
2	CAL2	1	Water	NA	NA	20.3	7.01	02/02/2022	17:01
3	CAL3	1	Water	NA	NA	20.3	10.03	02/02/2022	17:02
4	ICV	1	Water	NA	NA	20.4	6.98	02/02/2022	17:05
5	CCV1	1	Water	NA	NA	20.4	2.01	02/02/2022	17:08
6	N1366-01	1	Water	NA	NA	20.2	6.21	02/02/2022	17:10
7	N1368-01	1	Water	NA	NA	20.2	8.71	02/02/2022	17:13
8	N1383-01	1	Water	NA	NA	20.6	7.64	02/02/2022	17:15
9	N1383-01DUP	1	Water	NA	NA	20.7	7.65	02/02/2022	17:16
10	CCV2	1	Water	NA	NA	20.4	12.03	02/02/2022	17:19

WORKLIST(Hardcopy Internal Chain)

W 118500

WorkList Name : ph n1366 WorkList ID : 156522 Department : Wet-Chemistry Date : 02-02-2022 15:50:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/02/2022	Water	N1366-01	pH	Cool 4 deg C	PSEG03	M11	N28660	02/02/2022	9040C
02/02/2022	Water	N1368-01	pH	Cool 4 deg C	PSEG03	M11	31382	02/02/2022	9040C
02/02/2022	Water	N1383-01	pH	Cool 4 deg C	PSEG04	M11	402	02/02/2022	9040C

Date/Time 02/02/2022 17:130
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JB (WC)

Date/Time 02/02/2022 16:40
 Raw Sample Received by: JB (WC)
 Raw Sample Relinquished by: KS (WC)