

LB118503


7470A

INSTRUMENT ID: CVZ

Linear

 μ Abs.:

41389



A= 0.0000e+000

B= 2.4226e-004

C= -1.4719e-002

Rho= 0.9999940

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.016	0.016	127	0.000	127					
0.2	0.200	0.193	-0.007	856	0.0 %	856					-4
2.5	2.500	2.489	-0.011	10333	0.0 %	10333					0
5.0	5.000	5.009	0.009	20736	0.0 %	20736					0
7.5	7.500	7.482	-0.018	30946	0.0 %	30946					0
10.0	10.000	10.012	0.012	41389	0.0 %	41389					0

LB118503 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Feb 2022 10:48:23

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	127	-	0.0000	7470A	PPB	03 Feb 2022 11:05:46	S	Std
0.2 - 1	50.2	856	-	0.2000	7470A	PPB	03 Feb 2022 11:08:03	S	Std
2.5 - 1	52.5	10333	-	2.5000	7470A	PPB	03 Feb 2022 11:10:20	S	Std
5.0 - 1	55	20736	-	5.0000	7470A	PPB	03 Feb 2022 11:12:37	S	Std
7.5 - 1	57.5	30946	-	7.5000	7470A	PPB	03 Feb 2022 11:14:53	S	Std
10.0 - 1	510	41389	-	10.0000	7470A	PPB	03 Feb 2022 11:17:12	S	Std
ICV71 - 1	ICV71	16165	3.9013	-	7470A	PPB	03 Feb 2022 11:22:28	U	SMPL
ICB71 - 1	ICB71	0	-0.0147	-	7470A	PPB	03 Feb 2022 11:24:43	U	SMPL
CCV14 - 1	CCV14	20002	4.8309	-	7470A	PPB	03 Feb 2022 11:26:58	U	SMPL
CCB14 - 1	CCB14	33	-0.0067	-	7470A	PPB	03 Feb 2022 11:29:13	U	SMPL
CRA - 1	CRA	735	0.1633	-	7470A	PPB	03 Feb 2022 11:31:28	U	SMPL
HighStd - 1	HighStd	39648	9.5902	-	7470A	PPB	03 Feb 2022 11:33:43	U	SMPL
ChkStd - 1	ChkStd	27663	6.6868	-	7470A	PPB	03 Feb 2022 11:35:58	U	SMPL
PB142445BL - 1	PBW	-50	-0.0268	-	7470A	PPB	03 Feb 2022 11:38:14	U	SMPL
PB142445BS - 1	LCSW	14816	3.5745	-	7470A	PPB	03 Feb 2022 11:40:31	U	SMPL
N1357-02 - 1	SOIL-CUTTINGS	13	-0.0116	-	7470A	PPB	03 Feb 2022 11:42:47	U	SMPL
N1357-02DUP - 1	SOIL-CUTTINGSDUP	101	0.0097	-	7470A	PPB	03 Feb 2022 11:45:05	U	SMPL
N1357-02MS - 1	SOIL-CUTTINGSMS	14528	3.5048	-	7470A	PPB	03 Feb 2022 11:47:21	U	SMPL
N1357-02MSD - 1	SOIL-CUTTINGSMSD	14551	3.5103	-	7470A	PPB	03 Feb 2022 11:49:38	U	SMPL
PB142446BL - 1	PBW	-9	-0.0169	-	7470A	PPB	03 Feb 2022 11:51:54	U	SMPL
CCV15 - 1	CCV15	19782	4.7776	-	7470A	PPB	03 Feb 2022 11:54:09	U	SMPL
CCB15 - 1	CCB15	34	-0.0065	-	7470A	PPB	03 Feb 2022 11:56:25	U	SMPL
PB142446BS - 1	LCSW	14961	3.6097	-	7470A	PPB	03 Feb 2022 11:58:40	U	SMPL
N1357-01 - 1	GROUNDWATER	297	0.0572	-	7470A	PPB	03 Feb 2022 12:05:46	U	SMPL
N1357-01DUP - 1	GROUNDWATERDUP	337	0.0669	-	7470A	PPB	03 Feb 2022 12:08:01	U	SMPL
N1357-01MS - 1	GROUNDWATERMS	15236	3.6763	-	7470A	PPB	03 Feb 2022 12:12:50	U	SMPL
N1357-01MSD - 1	GROUNDWATERMSD	15197	3.6668	-	7470A	PPB	03 Feb 2022 12:15:06	U	SMPL
PB142420TB - 1	PB142420TB	0	-0.0147	-	7470A	PPB	03 Feb 2022 12:17:23	U	SMPL
N1357-02LX5 - 1		12	-0.0118	-	7470A	PPB	03 Feb 2022 12:21:39	U	SMPL
N1357-02A - 1		15661	3.7792	-	7470A	PPB	03 Feb 2022 12:23:55	U	SMPL
N1357-01LX - 1		51	-0.0024	-	7470A	PPB	03 Feb 2022 12:26:12	U	SMPL
N1357-01A - 1		15824	3.8187	-	7470A	PPB	03 Feb 2022 12:28:28	U	SMPL
CCV16 - 1	CCV16	19994	4.8289	-	7470A	PPB	03 Feb 2022 12:30:44	U	SMPL
CCB16 - 1	CCB16	33	-0.0067	-	7470A	PPB	03 Feb 2022 12:32:59	U	SMPL