



Analytical Summary Report

Analysis Method: SM2120 B

Reviewed By: Iwona

Parameter: Color

Supervisor Review By: apatel

Run Number: LB118506

pH Meter ID: WC pH METER-1

Reagent/Standard	Lot/Log #
Calibration std, 0 color unit	WP94763
Calibration std, 5 color unit	WP94764
Calibration std, 10 color unit	WP94765
Calibration std, 15 color unit	WP94766
Calibration std, 20 color unit	WP94767
Calibration std, 25 color unit	WP94768
Calibration std, 30 color unit	WP94769
Calibration std, 40 color unit	WP94770
Calibration std, 50 color unit	WP94771
Calibration std, 60 color unit	WP94772
Calibration std, 70 color unit	WP94773

Seq	LabID	ClientID	DL	pH Reading	pH Adjustment	Result (cu)	Anal Date	Anal Time
1	LB118506BL	LB118506BL	1	5.08	N/A	0.00	02/03/2022	10:40
2	N1375-01	MONTHLY-OUTFALL-002	20	8.26	N/A	40.00	02/03/2022	10:45
3	N1375-01DUP	MONTHLY-OUTFALL-002	20	8.22	N/A	40.00	02/03/2022	10:51