



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118507

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	02/03/2022	11:15
2	CAL2	1	Water	NA	NA	20.5	7.00	02/03/2022	11:16
3	CAL3	1	Water	NA	NA	20.5	10.03	02/03/2022	11:17
4	ICV	1	Water	NA	NA	20.6	6.98	02/03/2022	11:19
5	CCV1	1	Water	NA	NA	20.4	2.01	02/03/2022	11:20
6	N1370-01	1	Water	NA	NA	20.6	8.54	02/03/2022	11:21
7	N1370-01DUP	1	Water	NA	NA	20.7	8.55	02/03/2022	11:22
8	N1372-01	1	Water	NA	NA	20.1	8.54	02/03/2022	11:24
9	CCV2	1	Water	NA	NA	20.6	12.03	02/03/2022	11:25

LB 118507

WORKLIST(Hardcopy Internal Chain)

WorkList Name: PH N1370 WorkList ID: 156545 Department: Wet-Chemistry Date: 02-03-2022 11:08:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/02/2022	Water	N1370-01	pH C	Cool 4 deg C	INFR01	N11	BI-WEEKLY-DECANT	02/02/2022	9040C
02/02/2022	Water	N1372-01	pH D	Cool 4 deg C	INFR01	N11	BI-WEEKLY-INFLUENT	02/02/2022	9040C

Date/Time 02/03/22 11:10
Raw Sample Received by: 40 (WC)
Raw Sample Relinquished by: C. J. SM

Date/Time 02/03/22 12:30
Raw Sample Received by: C. J. SM
Raw Sample Relinquished by: 40 (WC)