



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 2/4/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 02/03/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:27
Out Date: 02/04/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB118508

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1381-01	DB-05 (20-22) -02-01-22	1	1.12	8.72	9.84	9.05	90.9	
N1381-02	DB-05 (33-35) -02-01-22	2	1.18	8.70	9.88	8.3	81.8	
N1381-03	DB-06 (15-20) -02-01-22	3	1.15	8.80	9.95	8.43	82.7	
N1381-04	DB-06 (38-40) -02-01-22	4	1.16	8.81	9.97	8.88	87.6	
N1381-05	DB-11 (15-20) -02-02-22	5	1.16	8.48	9.64	8.45	86.0	
N1381-06	DB-11 (32-34) -02-02-22	6	1.14	8.80	9.94	8.29	81.3	
N1381-07	DB-07 (13.5-15) -02-02-22	7	1.19	8.64	9.83	8.81	88.2	
N1381-08	DB-07 (20-22) -02-02-22	8	1.15	8.78	9.93	8.73	86.3	
N1381-09	DB-09 (10-15) -02-02-22	9	1.18	8.73	9.91	8.93	88.8	
N1381-10	DB-09 (25-30) -02-02-22	10	1.1	8.79	9.89	8.33	82.3	
N1381-11	N1381-10MS	11	1.1	8.79	9.89	8.33	82.3	
N1381-12	N1381-10MSD	12	1.1	8.79	9.89	8.33	82.3	
N1384-01	26982	13	1.17	8.42	9.59	7.84	79.2	
N1384-02	26982-EPH-2	14	1.17	8.67	9.84	8.3	82.2	
N1387-01	26516	15	1.15	8.49	9.64	7.5	74.8	
N1387-02	26516-EPH-2	16	1.12	8.71	9.83	7.77	76.3	
N1387-03	27116	17	1.19	8.55	9.74	8.82	89.2	
N1387-04	27116-EPH-2	18	1.19	8.53	9.72	8.78	89.0	
N1387-05	26601	19	1.14	8.51	9.65	8.46	86.0	
N1387-06	26601-EPH-2	20	1.13	8.69	9.82	8.69	87.0	
N1387-07	24801	21	1.15	8.80	9.95	8.94	88.5	
N1387-08	24801-EPH-2	22	1.17	8.50	9.67	8.89	90.8	
N1391-01	TR-04-020322	23	1.11	8.55	9.66	8.52	86.7	
N1391-02	TR-04-020322-EPH-2	24	1.11	8.81	9.92	9.19	91.7	
N1391-04	TR-08-020322	25	1.16	8.80	9.96	8.54	83.9	
N1391-05	TR-08-020322-EPH-2	26	1.19	8.79	9.98	8.8	86.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$