


LB118512

7471B

INSTRUMENT ID: CVL



Linear

A= 0.0000e+000

B= 1.9291e-004

C= 1.0599e-002

Rho= 0.9999761

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.021	0.021	53	0.000	53					10
0.2	0.200	0.200	0.000	984	0.0 %	984					0
2.5	2.500	2.508	0.008	12946	0.0 %	12946					0
5.0	5.000	4.948	-0.052	25593	0.0 %	25593					-10
7.5	7.500	7.497	-0.003	38809	0.0 %	38809					0
10.0	10.000	10.026	0.026	51917	0.0 %	51917					0

LB118512 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Feb 2022 14:03:28

Sample ID	Extended ID	µ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	SD	53	-	0.00007471B	PPB	03 Feb 2022 14:20:15	S	Std
0.2 - 1	SD	984	-	0.20007471B	PPB	03 Feb 2022 14:22:32	S	Std
2.5 - 1	SD	12946	-	2.50007471B	PPB	03 Feb 2022 14:24:50	S	Std
5.0 - 1	SD	25593	-	5.00007471B	PPB	03 Feb 2022 14:27:07	S	Std
7.5 - 1	SD	38809	-	7.50007471B	PPB	03 Feb 2022 14:29:24	S	Std
10.0 - 1	SD	51917	-	10.00007471B	PPB	03 Feb 2022 14:34:48	S	Std
ICV72 - 1	ICV72	20347	3.9357	-7471B	PPB	03 Feb 2022 14:37:32	U	SMPL
ICB72 - 1	ICB72	-14	0.0079	-7471B	PPB	03 Feb 2022 14:39:47	U	SMPL
CCV17 - 1	CCV17	26165	5.0581	-7471B	PPB	03 Feb 2022 14:42:02	U	SMPL
CCB17 - 1	CCB17	-36	0.0037	-7471B	PPB	03 Feb 2022 14:44:17	U	SMPL
CRA - 1	CRA	985	0.2006	-7471B	PPB	03 Feb 2022 14:46:33	U	SMPL
HighStd - 1	HighStd	52075	10.0563	-7471B	PPB	03 Feb 2022 14:48:47	U	SMPL
ChkStd - 1	ChkStd	36217	6.9972	-7471B	PPB	03 Feb 2022 14:51:02	U	SMPL
PB142462BL - 1	PB142462BL	-55	-0.0000	-7471B	PPB	03 Feb 2022 14:53:20	U	SMPL
PB142462BS - 1	PB142462BS	20396	3.9452	-7471B	PPB	03 Feb 2022 14:55:36	U	SMPL
N1362-01 - 1	RO-342	1835	0.3646	-7471B	PPB	03 Feb 2022 14:57:53	U	SMPL
N1362-03 - 1	RO-341	1371	0.2751	-7471B	PPB	03 Feb 2022 15:00:10	U	SMPL
N1367-01 - 1	31246	203	0.0498	-7471B	PPB	03 Feb 2022 15:02:26	U	SMPL
N1373-01 - 1	OR-3-020222	87123	16.8174	-7471B	PPB	03 Feb 2022 15:04:44	U	SMPL
N1378-01 - 1	ASTORIA-YARD-PIT-11-STOCKPILE	43806	8.4612	-7471B	PPB	03 Feb 2022 15:07:00	U	SMPL
CCV18 - 1	CCV18	26150	5.0552	-7471B	PPB	03 Feb 2022 15:09:18	U	SMPL
CCB18 - 1	CCB18	-34	0.0040	-7471B	PPB	03 Feb 2022 15:11:34	U	SMPL
N1382-04 - 1	WASTE-BUILDING	311	0.0706	-7471B	PPB	03 Feb 2022 15:13:50	U	SMPL
N1382-04DUP - 1	WASTE-BUILDINGDUP	347	0.0775	-7471B	PPB	03 Feb 2022 15:16:06	U	SMPL
N1382-04MS - 1	WASTE-BUILDINGMS	21782	4.2125	-7471B	PPB	03 Feb 2022 15:18:21	U	SMPL
N1382-04MSD - 1	WASTE-BUILDINGMSD	20342	3.9348	-7471B	PPB	03 Feb 2022 15:20:37	U	SMPL
N1382-04LX5 - 1		43	0.0189	-7471B	PPB	03 Feb 2022 15:22:53	U	SMPL
N1382-04A - 1		20523	3.9697	-7471B	PPB	03 Feb 2022 15:25:09	U	SMPL
N1373-01DLX5 - 1		16806	3.2526	-7471B	PPB	03 Feb 2022 15:27:25	U	SMPL
CCV19 - 1	CCV19	26145	5.0542	-7471B	PPB	03 Feb 2022 15:29:42	U	SMPL
CCB19 - 1	CCB19	-32	0.0044	-7471B	PPB	03 Feb 2022 15:31:59	U	SMPL
N1356-01 - 1	F-PHC	32942	6.3654	-7471B	PPB	03 Feb 2022 15:42:33	U	SMPL
N1356-01DUP - 1	F-PHCDUP	32453	6.2711	-7471B	PPB	03 Feb 2022 15:47:04	U	SMPL
N1356-01MS - 1	F-PHCMS	47540	9.1815	-7471B	PPB	03 Feb 2022 15:49:20	U	SMPL
N1356-01MSD - 1	F-PHCMSD	48085	9.2866	-7471B	PPB	03 Feb 2022 15:54:12	U	SMPL
N1356-01LX5 - 1		6426	1.2502	-7471B	PPB	03 Feb 2022 15:56:27	U	SMPL
N1356-01A - 1		51683	9.9807	-7471B	PPB	03 Feb 2022 15:58:44	U	SMPL
CCV20 - 1	CCV20	26191	5.0631	-7471B	PPB	03 Feb 2022 16:00:59	U	SMPL
CCB20 - 1	CCB20	-26	0.0056	-7471B	PPB	03 Feb 2022 16:03:16	U	SMPL