



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118516

Slope : 98.7

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/03/2022	15:30
2	CAL2	1	Water	NA	NA	20.2	7.00	02/03/2022	15:31
3	CAL3	1	Water	NA	NA	20.3	10.03	02/03/2022	15:33
4	ICV	1	Water	NA	NA	20.1	6.99	02/03/2022	15:34
5	CCV1	1	Water	NA	NA	20.5	2.01	02/03/2022	15:35
6	N1392-01	1	Water	NA	NA	20.1	6.50	02/03/2022	15:36
7	N1394-01	1	Water	NA	NA	20.3	4.87	02/03/2022	15:39
8	N1394-01DUP	1	Water	NA	NA	20.4	4.88	02/03/2022	15:40
9	CCV2	1	Water	NA	NA	20.4	12.03	02/03/2022	15:42

NP 118516

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1392 WorkList ID : 156555 Department : Wet-Chemistry Date : 02-03-2022 15:17:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/03/2022	Water	N1392-01	pH	Cool 4 deg C	PSEG03	M12	OILY-WATER-NWB-1735	02/03/2022	9040C
01/31/2022	Water	N1394-01	pH	Cool 4 deg C	STER11	N13	POOL-WATER	01/31/2022	9040C

Date/Time 02/03/22 15:26
Raw Sample Received by: J.P. (S.W.)
Raw Sample Relinquished by: J.P. (S.W.)

Date/Time 02/03/22 16:30
Raw Sample Received by: J.P. (S.W.)
Raw Sample Relinquished by: J.P. (S.W.)