



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118522

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	02/03/2022	17:15
2	CAL2	1	Water	NA	NA	20.8	7.00	02/03/2022	17:16
3	CAL3	1	Water	NA	NA	20.7	10.03	02/03/2022	17:18
4	ICV	1	Water	NA	NA	20.7	6.99	02/03/2022	17:20
5	CCV1	1	Water	NA	NA	20.7	2.01	02/03/2022	17:21
6	N1396-01	1	Water	NA	NA	20.6	7.08	02/03/2022	17:22
7	N1396-01DUP	1	Water	NA	NA	20.7	7.09	02/03/2022	17:23
8	N1396-03	1	Water	NA	NA	20.3	6.98	02/03/2022	17:25
9	N1396-05	1	Water	NA	NA	20.1	6.91	02/03/2022	17:26
10	CCV2	1	Water	NA	NA	20.7	12.03	02/03/2022	17:30

VP 118522

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1396 WorkList ID : 156561 Department : Wet-Chemistry Date : 02-03-2022 17:04:53

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Raw Sample Storage Location		
02/03/2022	Water	N1396-01	pH <i>A</i>	Cool 4 deg C	PSEG04	M11 DSN002	02/03/2022	9040C
02/03/2022	Water	N1396-03	pH <i>A</i>	Cool 4 deg C	PSEG04	M11 DSN001	02/03/2022	9040C
02/03/2022	Water	N1396-05	pH <i>A</i>	Cool 4 deg C	PSEG04	M11 RWS003	02/03/2022	9040C

Date/Time 02/03/22 17:10
Raw Sample Received by: JD CSM
Raw Sample Relinquished by: JD CSM

Date/Time 02/03/22 18:20
Raw Sample Received by: JD CSM
Raw Sample Relinquished by: JP CSM