

LB11852 F


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41215



A= 0.0000e+000

B= 2.4394e-004

C= 2.5876e-002

Rho= 0.9998688

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%SD
0.0	0.000	0.031	0.031	19	0.000	19					-
0.2	0.200	0.190	-0.010	674	0.0 %	674					-5
2.5	2.500	2.508	0.008	10176	0.0 %	10176					0
5.0	5.000	5.009	0.009	20427	0.0 %	20427					0
7.5	7.500	7.382	-0.118	30156	0.0 %	30156					-2
10.0	10.000	10.080	0.080	41215	0.0 %	41215					1

LB118525 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Feb 2022 10:54:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	19	-	0.0000	7470A	PPB	04 Feb 2022 11:10:20	S	Std
0.2 - 1	50.2	674	-	0.2000	7470A	PPB	04 Feb 2022 11:12:37	S	Std
2.5 - 1	52.5	10176	-	2.5000	7470A	PPB	04 Feb 2022 11:14:54	S	Std
5.0 - 1	55	20427	-	5.0000	7470A	PPB	04 Feb 2022 11:17:11	S	Std
7.5 - 1	57.5	30156	-	7.5000	7470A	PPB	04 Feb 2022 11:19:27	S	Std
10.0 - 1	510	41215	-	10.0000	7470A	PPB	04 Feb 2022 11:21:43	S	Std
ICV73 - 1	ICV73	15870	3.8972	-	7470A	PPB	04 Feb 2022 11:24:31	U	SMPL
ICB73 - 1	ICB73	-18	0.0215	-	7470A	PPB	04 Feb 2022 11:26:46	U	SMPL
CCV21 - 1	CCV21	19930	4.8876	-	7470A	PPB	04 Feb 2022 11:31:53	U	SMPL
CCB21 - 1	CCB21	-13	0.0227	-	7470A	PPB	04 Feb 2022 11:34:08	U	SMPL
CRA - 1	CRA	667	0.1886	-	7470A	PPB	04 Feb 2022 11:36:23	U	SMPL
HighStd - 1	HighStd	39833	9.7428	-	7470A	PPB	04 Feb 2022 11:38:38	U	SMPL
ChkStd - 1	ChkStd	28116	6.8845	-	7470A	PPB	04 Feb 2022 11:40:53	U	SMPL
PB142470BL - 1	PBW	-50	0.0137	-	7470A	PPB	04 Feb 2022 11:43:10	U	SMPL
PB142470BS - 1	LCSW	13997	3.4403	-	7470A	PPB	04 Feb 2022 11:45:27	U	SMPL
N1356-02 - 1	F-PHC	6	0.0273	-	7470A	PPB	04 Feb 2022 11:47:44	U	SMPL
N1356-02DUP - 1	F-PHCDUP	62	0.0410	-	7470A	PPB	04 Feb 2022 11:50:00	U	SMPL
N1356-02MS - 1	F-PHCMS	13975	3.4350	-	7470A	PPB	04 Feb 2022 11:52:17	U	SMPL
N1356-02MSD - 1	F-PHCMSD	14287	3.5111	-	7470A	PPB	04 Feb 2022 11:54:34	U	SMPL
N1356-04 - 1	M-PHC	-11	0.0232	-	7470A	PPB	04 Feb 2022 11:56:50	U	SMPL
CCV22 - 1	CCV22	19918	4.8847	-	7470A	PPB	04 Feb 2022 11:59:05	U	SMPL
CCB22 - 1	CCB22	-23	0.0203	-	7470A	PPB	04 Feb 2022 12:01:20	U	SMPL
N1356-06 - 1	B-PHC	25	0.0320	-	7470A	PPB	04 Feb 2022 12:03:37	U	SMPL
PB142471BL - 1	PBW	27	0.0325	-	7470A	PPB	04 Feb 2022 12:05:53	U	SMPL
PB142471BS - 1	LCSW	13610	3.3459	-	7470A	PPB	04 Feb 2022 12:08:08	U	SMPL
N1360-02 - 1	COMP	481	0.1432	-	7470A	PPB	04 Feb 2022 12:10:24	U	SMPL
N1366-01 - 1	N28660	302	0.0995	-	7470A	PPB	04 Feb 2022 12:12:40	U	SMPL
N1366-01DUP - 1	N28660DUP	275	0.0930	-	7470A	PPB	04 Feb 2022 12:15:42	U	SMPL
N1366-01MS - 1	N28660MS	15416	3.7865	-	7470A	PPB	04 Feb 2022 12:20:37	U	SMPL
N1366-01MSD - 1	N28660MSD	15398	3.7821	-	7470A	PPB	04 Feb 2022 12:22:53	U	SMPL
PB142439TB - 1	PB142439TB	-21	0.0208	-	7470A	PPB	04 Feb 2022 12:25:10	U	SMPL
CCV23 - 1	CCV23	19924	4.8862	-	7470A	PPB	04 Feb 2022 12:27:26	U	SMPL
CCB23 - 1	CCB23	-23	0.0203	-	7470A	PPB	04 Feb 2022 12:29:41	U	SMPL
N1356-02LX5 - 1		10	0.0283	-	7470A	PPB	04 Feb 2022 12:31:57	U	SMPL
N1356-02A - 1		15477	3.8014	-	7470A	PPB	04 Feb 2022 12:34:13	U	SMPL
N1366-01LX5 - 1		102	0.0508	-	7470A	PPB	04 Feb 2022 12:36:29	U	SMPL
N1366-01A - 1		15629	3.8384	-	7470A	PPB	04 Feb 2022 12:38:45	U	SMPL
CCV24 - 1	CCV24	19713	4.8347	-	7470A	PPB	04 Feb 2022 12:41:00	U	SMPL
CCB24 - 1	CCB24	-34	0.0176	-	7470A	PPB	04 Feb 2022 12:43:17	U	SMPL