



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118532

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/04/2022	12:16
2	CAL2	1	Water	NA	NA	20.3	7.00	02/04/2022	12:17
3	CAL3	1	Water	NA	NA	20.4	10.03	02/04/2022	12:19
4	ICV	1	Water	NA	NA	20.3	6.99	02/04/2022	12:20
5	CCV1	1	Water	NA	NA	20.3	2.01	02/04/2022	12:21
6	N1400-02	1	Water	NA	NA	20.6	6.73	02/04/2022	12:25
7	N1404-01	1	Water	NA	NA	20.5	4.90	02/04/2022	12:26
8	N1404-01DUP	1	Water	NA	NA	20.6	4.91	02/04/2022	12:27
9	CCV2	1	Water	NA	NA	20.6	12.03	02/04/2022	12:30

LB 118532

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1404 WorkList ID : 156585 Department : Wet-Chemistry Date : 02-04-2022 12:11:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Customer Sample	Collect Date	Method
						Storage	Location			
02/03/2022	Water	N1400-02	pH	Cool 4 deg C	PARS15	M12	KTS1-WC-L-MH-02		02/03/2022	9040C
01/29/2022	Water	N1404-01	pH	Cool 4 deg C	STER15	N21	POOL-WATER		01/29/2022	9040C

Date/Time 02/04/22 12:15
Raw Sample Received by: JP (WCI)
Raw Sample Relinquished by: JP SM

Date/Time 02/04/22 13:30
Raw Sample Received by: JP SM
Raw Sample Relinquished by: JP (WCI)