

LB118533


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

53114



A= 0.0000e+000

B= 1.8789e-004

C= -5.9418e-003

Rho= 0.9999797

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.001	0.001	38	0.000	38					
0.2	0.200	0.166	-0.034	915	0.0 %	915					
2.5	2.500	2.531	0.031	13504	0.0 %	13504					
5.0	5.000	5.008	0.008	26686	0.0 %	26686					
7.5	7.500	7.520	0.020	40057	0.0 %	40057					
10.0	10.000	9.973	-0.027	53114	0.0 %	53114					

1001
-17
1
0
0
0

LB118533 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 04 Feb 2022 13:23:00					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	38	-	0.0000	7471B	PPB	04 Feb 2022 13:29:30	S	Std
0.2 - 1	50.2	915	-	0.2000	7471B	PPB	04 Feb 2022 13:31:47	S	Std
2.5 - 1	52.5	13504	-	2.5000	7471B	PPB	04 Feb 2022 13:34:04	S	Std
5.0 - 1	55	26686	-	5.0000	7471B	PPB	04 Feb 2022 13:36:21	S	Std
7.5 - 1	57.5	40057	-	7.5000	7471B	PPB	04 Feb 2022 13:38:37	S	Std
10.0 - 1	510	53114	-	10.0000	7471B	PPB	04 Feb 2022 13:40:53	S	Std
ICV74 - 1	ICV74	20592	3.8630	-	7471B	PPB	04 Feb 2022 13:43:40	U	SMPL
ICB74 - 1	ICB74	-44	-0.0142	-	7471B	PPB	04 Feb 2022 13:45:55	U	SMPL
CCV25 - 1	CCV25	26201	4.9168	-	7471B	PPB	04 Feb 2022 13:48:10	U	SMPL
CCB25 - 1	CCB25	136	0.0196	-	7471B	PPB	04 Feb 2022 13:50:25	U	SMPL
CRA - 1	CRA	884	0.1601	-	7471B	PPB	04 Feb 2022 13:52:41	U	SMPL
HighStd - 1	HighStd	52081	9.7793	-	7471B	PPB	04 Feb 2022 13:54:56	U	SMPL
ChkStd - 1	ChkStd	36508	6.8534	-	7471B	PPB	04 Feb 2022 13:57:12	U	SMPL
PB142494BL - 1	PBS	-77	-0.0204	-	7471B	PPB	04 Feb 2022 13:59:29	U	SMPL
PB142494BS - 1	LCSS	18502	3.4703	-	7471B	PPB	04 Feb 2022 14:01:47	U	SMPL
N1384-01 - 1	26982	5620	1.0500	-	7471B	PPB	04 Feb 2022 14:04:03	U	SMPL
N1387-01 - 1	26516	21834	4.0963	-	7471B	PPB	04 Feb 2022 14:06:19	U	SMPL
N1387-03 - 1	27116	3948	0.7358	-	7471B	PPB	04 Feb 2022 14:08:36	U	SMPL
N1387-05 - 1	26601	15610	2.9270	-	7471B	PPB	04 Feb 2022 14:10:53	U	SMPL
N1387-07 - 1	24801	1889	0.3490	-	7471B	PPB	04 Feb 2022 14:13:08	U	SMPL
CCV26 - 1	CCV26	26534	4.9794	-	7471B	PPB	04 Feb 2022 14:15:24	U	SMPL
CCB26 - 1	CCB26	-42	-0.0138	-	7471B	PPB	04 Feb 2022 14:17:39	U	SMPL
N1391-01 - 1	TR-04-020322	976	0.1774	-	7471B	PPB	04 Feb 2022 14:19:55	U	SMPL
N1391-01DUP - 1	TR-04-020322DUP	991	0.1803	-	7471B	PPB	04 Feb 2022 14:22:11	U	SMPL
N1391-01MS - 1	TR-04-020322MS	18763	3.5194	-	7471B	PPB	04 Feb 2022 14:24:27	U	SMPL
N1391-01MSD - 1	TR-04-020322MSD	18771	3.5209	-	7471B	PPB	04 Feb 2022 14:26:42	U	SMPL
N1391-04 - 1	TR-08-020322	795	0.1434	-	7471B	PPB	04 Feb 2022 14:28:58	U	SMPL
N1391-01LX5 - 1		179	0.0277	-	7471B	PPB	04 Feb 2022 14:31:14	U	SMPL
N1391-01A - 1		20740	3.8908	-	7471B	PPB	04 Feb 2022 14:33:31	U	SMPL
CCV27 - 1	CCV27	26342	4.9433	-	7471B	PPB	04 Feb 2022 14:35:49	U	SMPL
CCB27 - 1	CCB27	106	0.0140	-	7471B	PPB	04 Feb 2022 14:38:06	U	SMPL