



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 2/7/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 15:30
In Date: 02/04/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 02/05/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB118535

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1408-01	A5700	1	1.12	8.78	9.9	8.92	88.8	
N1408-02	A5701	2	1.12	8.77	9.89	8.64	85.7	
N1408-03	A5702	3	1.12	8.83	9.95	8.98	89.0	
N1408-04	A5703	4	1.12	8.75	9.87	9.2	92.3	
N1408-05	A5704	5	1.12	8.78	9.9	8.75	86.9	
N1408-06	A5705	6	1.12	8.68	9.8	9.13	92.3	
N1408-07	A5706	7	1.12	8.73	9.85	9.09	91.3	
N1408-08	A5707	8	1.12	8.68	9.8	8.99	90.7	
N1408-09	A5708	9	1.12	8.83	9.95	9.1	90.4	
N1408-10	A5709	10	1.12	8.68	9.8	9.04	91.2	
N1408-11	A5710	11	1.12	8.77	9.89	9.22	92.4	
N1408-12	A5710D	12	1.12	8.77	9.89	9.22	92.4	
N1408-13	A5710S	13	1.12	8.77	9.89	9.22	92.4	
N1408-14	A5711	14	1.12	8.77	9.89	9.00	89.9	
N1408-15	A5712	15	1.09	8.85	9.94	9.16	91.2	
N1408-16	A5713	16	1.08	8.80	9.88	9.44	95.0	
N1408-17	A5714	17	1.07	8.79	9.86	8.95	89.6	
N1408-18	A5715	18	1.1	8.80	9.9	9.06	90.5	
N1408-19	A5716	19	1.08	8.90	9.98	8.74	86.1	
N1408-20	A5717	20	1.07	8.85	9.92	9.08	90.5	
N1408-21	A5718	21	1.12	8.73	9.85	8.86	88.7	
N1408-22	A5719	22	1.14	8.84	9.98	8.81	86.8	
N1408-23	A5733	23	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$