

LB118540


7470A

INSTRUMENT ID: CVK

Linear

 μ Abs.:

40910



A= 0.0000e+000

B= 2.4541e-004

C= 2.4389e-003

Rho= 0.9998480

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.013	0.013	45	0.000	45				
0.2	0.200	0.182	-0.018	732	0.0 %	732				
2.5	2.500	2.485	-0.015	10117	0.0 %	10117				
5.0	5.000	5.094	0.094	20745	0.0 %	20745				
7.5	7.500	7.383	-0.117	30076	0.0 %	30076				
10.0	10.000	10.042	0.042	40910	0.0 %	40910				

0/20

1
-9
-1
2
-2
0

LB118540 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Feb 2022 15:32:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	45	-	0.00007470A	PPB		04 Feb 2022 15:39:46	S	Std
0.2 - 1	50.2	732	-	0.20007470A	PPB		04 Feb 2022 15:42:04	S	Std
2.5 - 1	52.5	10117	-	2.50007470A	PPB		04 Feb 2022 15:44:21	S	Std
5.0 - 1	55	20745	-	5.00007470A	PPB		04 Feb 2022 15:46:38	S	Std
7.5 - 1	57.5	30076	-	7.50007470A	PPB		04 Feb 2022 15:48:55	S	Std
10.0 - 1	58	40910	-	10.00007470A	PPB		04 Feb 2022 15:53:41	S	Std
ICV75 - 1	ICV75	16389	4.0245	-7470A	PPB		04 Feb 2022 15:58:51	U	SMPL
ICB75 - 1	ICB75	-31	-0.0052	-7470A	PPB		04 Feb 2022 16:01:06	U	SMPL
CCV28 - 1	CCV28	19229	4.7215	-7470A	PPB		04 Feb 2022 16:03:21	U	SMPL
CCB28 - 1	CCB28	-49	-0.0096	-7470A	PPB		04 Feb 2022 16:05:36	U	SMPL
CRA - 1	CRA	567	0.1416	-7470A	PPB		04 Feb 2022 16:07:51	U	SMPL
HighStd - 1	HighStd	40556	9.9554	-7470A	PPB		04 Feb 2022 16:10:06	U	SMPL
ChkStd - 1	ChkStd	28527	7.0033	-7470A	PPB		04 Feb 2022 16:12:22	U	SMPL
PB142495BL - 1	PBW	-54	-0.0108	-7470A	PPB		04 Feb 2022 16:14:37	U	SMPL
PB142495BS - 1	LCSW	14555	3.5744	-7470A	PPB		04 Feb 2022 16:16:53	U	SMPL
N1392-01 - 1	OILY-WATER-NWB-1735	-21	-0.0027	-7470A	PPB		04 Feb 2022 16:19:10	U	SMPL
N1392-01DUP - 1	OILY-WATER-NWB-1735DUP	26	0.0088	-7470A	PPB		04 Feb 2022 16:21:26	U	SMPL
N1392-01MS - 1	OILY-WATER-NWB-1735MS	14615	3.5891	-7470A	PPB		04 Feb 2022 16:23:43	U	SMPL
N1392-01MSD - 1	OILY-WATER-NWB-1735MSD	15277	3.7516	-7470A	PPB		04 Feb 2022 16:26:00	U	SMPL
PB142477TB - 1	PB142477TB	-17	-0.0017	-7470A	PPB		04 Feb 2022 16:28:15	U	SMPL
CCV29 - 1	CCV29	19483	4.7838	-7470A	PPB		04 Feb 2022 16:30:31	U	SMPL
CCB29 - 1	CCB29	-47	-0.0091	-7470A	PPB		04 Feb 2022 16:32:46	U	SMPL
N1392-01LX5 - 1		17	0.0066	-7470A	PPB		04 Feb 2022 16:35:01	U	SMPL
N1392-01A - 1		16877	4.1443	-7470A	PPB		04 Feb 2022 16:37:16	U	SMPL
CCV30 - 1	CCV30	18903	4.6415	-7470A	PPB		04 Feb 2022 16:44:16	U	SMPL
CCB30 - 1	CCB30	-36	-0.0064	-7470A	PPB		04 Feb 2022 16:46:31	U	SMPL