



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118546

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/04/2022	17:10
2	CAL2	1	Water	NA	NA	20.3	7.00	02/04/2022	17:11
3	CAL3	1	Water	NA	NA	20.3	10.03	02/04/2022	17:12
4	ICV	1	Water	NA	NA	20.4	6.98	02/04/2022	17:15
5	CCV1	1	Water	NA	NA	20.3	2.01	02/04/2022	17:19
6	N1405-03	1	Water	NA	NA	20.3	7.25	02/04/2022	17:20
7	N1405-03DUP	1	Water	NA	NA	20.4	7.26	02/04/2022	17:21
8	N1405-04	1	Water	NA	NA	20.5	8.21	02/04/2022	17:23
9	N1405-07	1	Water	NA	NA	20.1	7.83	02/04/2022	17:25
10	CCV2	1	Water	NA	NA	20.3	12.03	02/04/2022	17:26

VB 118546

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1405 W WorkList ID : 156597 Department : Wet-Chemistry Date : 02-04-2022 16:52:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
				Cool 4 deg C	PSEG03	Storage	Location		
02/04/2022	Water	N1405-03	pH	Cool 4 deg C	PSEG03	M12	143	02/04/2022	9040C
02/04/2022	Water	N1405-04	pH	Cool 4 deg C	PSEG03	M12	144	02/04/2022	9040C
02/04/2022	Water	N1405-07	pH	Cool 4 deg C	PSEG03	M12	150	02/04/2022	9040C

Date/Time 02/04/2022 17:10:00
Raw Sample Received by: JP (WCC)
Raw Sample Relinquished by: RJ (EXT-1006)

Date/Time 02/04/2022 18:10
Raw Sample Received by: RJ (EXT-1006)
Raw Sample Relinquished by: JP (WCC)