



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118547

Slope : 98.8

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/04/2022	17:10
2	CAL2	1	Water	NA	NA	20.3	7.00	02/04/2022	17:11
3	CAL3	1	Water	NA	NA	20.3	10.03	02/04/2022	17:12
4	ICV	1	Water	NA	NA	20.4	6.98	02/04/2022	17:15
5	CCV1	1	Water	NA	NA	20.3	2.01	02/04/2022	17:19
6	N1405-01	1	Solid	20.03	20	20.6	7.70	02/04/2022	17:35
7	N1419-01	1	Solid	20.02	20	20.4	6.31	02/04/2022	17:36
8	N1419-01DUP	1	Solid	20.03	20	20.5	6.32	02/04/2022	17:37
9	CCV2	1	Water	NA	NA	20.4	12.03	02/04/2022	17:39

VB 118547

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PHS N1405 WorkList ID : 156598 Department : Wet-Chemistry Date : 02-04-2022 16:52:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/04/2022	Solid	N1405-01	pH	Cool 4 deg C	PSEG03	M12	148	02/04/2022	9045D
02/03/2022	Solid	N1419-01	pH	Cool 4 deg C	TULL01	N21	GRIT-STOCKPILE	02/03/2022	9045D

Date/Time 02/04/22 17:00
Raw Sample Received by: TA Calc
Raw Sample Relinquished by: RSC (Est-Rel)

Date/Time 02/04/22 18:10
Raw Sample Received by: RSC (Est-Rel)
Raw Sample Relinquished by: TA Calc