

LB 118551


SFAM01.1

INSTRUMENT ID: CV2

Linear

 $\mu$  Abs.:

41380



A= 0.0000e+000

B= 2.4187e-004

C= -2.3742e-002

Rho= 0.9999931

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.00   | 0.000  | -0.011 | -0.011 | 51    | 0.000      | 51    |       |       |       |       |    |
| 0.05   | 0.050  |        |        |       |            | 205   | 0     |       |       |       |    |
| 0.20   | 0.200  | 0.207  | 0.007  | 952   | 0.0 %      | 952   |       |       |       |       |    |
| 2.50   | 2.500  | 2.489  | -0.011 | 10388 | 0.0 %      | 10388 |       |       |       |       |    |
| 5.00   | 5.000  | 5.023  | 0.023  | 20865 | 0.0 %      | 20865 |       |       |       |       |    |
| 7.50   | 7.500  | 7.509  | 0.009  | 31142 | 0.0 %      | 31142 |       |       |       |       |    |
| 10.0   | 10.000 | 9.985  | -0.015 | 41380 | 0.0 %      | 41380 |       |       |       |       |    |

0/0 D  
1  
4  
0  
0  
0  
0

LB118551  
INSTRUMENT ID : CV1

| Sample ID  | Extended ID | $\mu$ Abs. | Conc.     | Std Conc | Method       | Units | Date           | Type |
|------------|-------------|------------|-----------|----------|--------------|-------|----------------|------|
|            | 0 S0        | 51         | -         |          | 0 SFAM01.1   | PPB   | 2/5/2022 11:31 | Std  |
|            | 0.2 S01     | 952        | -         |          | 0.2 SFAM01.1 | PPB   | 2/5/2022 11:34 | Std  |
|            | 2.5 S02     | 10388      | -         |          | 2.5 SFAM01.1 | PPB   | 2/5/2022 11:38 | Std  |
|            | 5 S03       | 20865      | -         |          | 5 SFAM01.1   | PPB   | 2/5/2022 11:40 | Std  |
|            | 7.5 S04     | 31142      | -         |          | 7.5 SFAM01.1 | PPB   | 2/5/2022 11:42 | Std  |
|            | 10 S05      | 41380      | -         |          | 10 SFAM01.1  | PPB   | 2/5/2022 11:45 | Std  |
| ICV011     | ICV011      | 15808      | 3.7997 -  |          | SFAM01.1     | PPB   | 2/5/2022 11:47 | SMPL |
| ICB011     | ICB011      | -18        | -0.0281 - |          | SFAM01.1     | PPB   | 2/5/2022 11:50 | SMPL |
| CCV031     | CCV031      | 20915      | 5.0349 -  |          | SFAM01.1     | PPB   | 2/5/2022 11:52 | SMPL |
| CCB031     | CCB031      | -25        | -0.0298 - |          | SFAM01.1     | PPB   | 2/5/2022 11:54 | SMPL |
| PB142505BL | PBW505      | 18         | -0.0194 - |          | SFAM01.1     | PPB   | 2/5/2022 11:56 | SMPL |
| N1304-03   | COBX6       | -9         | -0.0259 - |          | SFAM01.1     | PPB   | 2/5/2022 12:01 | SMPL |
| N1304-04   | COBX8       | 31         | -0.0162 - |          | SFAM01.1     | PPB   | 2/5/2022 12:03 | SMPL |
| N1304-05   | COBY0       | 32         | -0.016 -  |          | SFAM01.1     | PPB   | 2/5/2022 12:05 | SMPL |
| N1304-06   | COBY3       | 20         | -0.0189 - |          | SFAM01.1     | PPB   | 2/5/2022 12:08 | SMPL |
| N1304-07   | COBY7       | 17         | -0.0196 - |          | SFAM01.1     | PPB   | 2/5/2022 12:10 | SMPL |
| N1304-08   | COBY7D      | 27         | -0.0172 - |          | SFAM01.1     | PPB   | 2/5/2022 12:12 | SMPL |
| N1304-09   | COBY7S      | 4132       | 0.9757 -  |          | SFAM01.1     | PPB   | 2/5/2022 12:15 | SMPL |
| PB142506BL | PBW506      | 14         | -0.0204 - |          | SFAM01.1     | PPB   | 2/5/2022 12:17 | SMPL |
| N1306-01   | COBX7       | 26         | -0.0175 - |          | SFAM01.1     | PPB   | 2/5/2022 12:19 | SMPL |
| N1306-02   | COBX9       | 26         | -0.0175 - |          | SFAM01.1     | PPB   | 2/5/2022 12:21 | SMPL |
| N1306-03   | COBY1       | 34         | -0.0155 - |          | SFAM01.1     | PPB   | 2/5/2022 12:24 | SMPL |
| N1306-04   | COBY4       | 35         | -0.0153 - |          | SFAM01.1     | PPB   | 2/5/2022 12:26 | SMPL |
| N1306-05   | COBY8       | 35         | -0.0153 - |          | SFAM01.1     | PPB   | 2/5/2022 12:28 | SMPL |
| N1306-06   | COBY8D      | 31         | -0.0162 - |          | SFAM01.1     | PPB   | 2/5/2022 12:30 | SMPL |
| N1306-07   | COBY8S      | 4732       | 1.1208 -  |          | SFAM01.1     | PPB   | 2/5/2022 12:33 | SMPL |
| N1306-08   | COBZ0       | 15         | -0.0201 - |          | SFAM01.1     | PPB   | 2/5/2022 12:35 | SMPL |
| PB142507BL | PBW507      | 23         | -0.0182 - |          | SFAM01.1     | PPB   | 2/5/2022 12:37 | SMPL |
| N1329-01   | COBT8       | 32         | -0.016 -  |          | SFAM01.1     | PPB   | 2/5/2022 12:39 | SMPL |
| CCV032     | CCV032      | 20713      | 4.9861 -  |          | SFAM01.1     | PPB   | 2/5/2022 12:42 | SMPL |
| CCB032     | CCB032      | -20        | -0.0286 - |          | SFAM01.1     | PPB   | 2/5/2022 12:44 | SMPL |
| N1329-02   | COBW0       | 48         | -0.0121 - |          | SFAM01.1     | PPB   | 2/5/2022 12:46 | SMPL |
| N1329-03   | COBW1       | 19         | -0.0191 - |          | SFAM01.1     | PPB   | 2/5/2022 12:49 | SMPL |
| N1329-04   | COBW4       | 22         | -0.0184 - |          | SFAM01.1     | PPB   | 2/5/2022 12:51 | SMPL |
| N1329-05   | COBW4D      | 24         | -0.0179 - |          | SFAM01.1     | PPB   | 2/5/2022 12:53 | SMPL |
| N1329-06   | COBW4S      | 3677       | 0.8656 -  |          | SFAM01.1     | PPB   | 2/5/2022 12:55 | SMPL |
| N1329-07   | COBW7       | 12         | -0.0208 - |          | SFAM01.1     | PPB   | 2/5/2022 12:58 | SMPL |
| N1329-08   | COBX0       | 45         | -0.0129 - |          | SFAM01.1     | PPB   | 2/5/2022 13:00 | SMPL |
| N1329-09   | COBX2       | 39         | -0.0143 - |          | SFAM01.1     | PPB   | 2/5/2022 13:02 | SMPL |
| PB142508BL | PBW508      | 26         | -0.0175 - |          | SFAM01.1     | PPB   | 2/5/2022 13:04 | SMPL |
| N1330-01   | COBT9       | 33         | -0.0158 - |          | SFAM01.1     | PPB   | 2/5/2022 13:07 | SMPL |
| N1330-02   | COBW2       | 30         | -0.0165 - |          | SFAM01.1     | PPB   | 2/5/2022 13:09 | SMPL |
| N1330-03   | COBW3       | 18         | -0.0194 - |          | SFAM01.1     | PPB   | 2/5/2022 13:11 | SMPL |
| N1330-04   | COBW5       | 23         | -0.0182 - |          | SFAM01.1     | PPB   | 2/5/2022 13:13 | SMPL |
| N1330-05   | COBW5D      | 18         | -0.0194 - |          | SFAM01.1     | PPB   | 2/5/2022 13:16 | SMPL |
| N1330-06   | COBW5S      | 3722       | 0.8765 -  |          | SFAM01.1     | PPB   | 2/5/2022 13:18 | SMPL |

LB118551

INSTRUMENT ID : CV1

|          |        |       |           |          |     |                     |
|----------|--------|-------|-----------|----------|-----|---------------------|
| N1330-07 | COBW8  | 12    | -0.0208 - | SFAM01.1 | PPB | 2/5/2022 13:20 SMPL |
| N1330-08 | COBX1  | 24    | -0.0179 - | SFAM01.1 | PPB | 2/5/2022 13:23 SMPL |
| N1330-09 | COBX3  | 24    | -0.0179 - | SFAM01.1 | PPB | 2/5/2022 13:25 SMPL |
| N1330-10 | COBX5  | 29    | -0.0167 - | SFAM01.1 | PPB | 2/5/2022 13:27 SMPL |
| N1304-01 | COBT2  | 19107 | 4.5976 -  | SFAM01.1 | PPB | 2/5/2022 13:29 SMPL |
| CCV033   | CCV033 | 20391 | 4.9082 -  | SFAM01.1 | PPB | 2/5/2022 13:34 SMPL |
| CCB033   | CCB033 | -26   | -0.03 -   | SFAM01.1 | PPB | 2/5/2022 13:36 SMPL |