

LB118552


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

52839



A= 0.0000e+000

B= 1.8893e-004

C= -1.0345e-002

Rho= 0.9999260

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>1/SD</i>
0.0	0.000	-0.007	-0.007	17	0.000	17					
0.2	0.200	0.199	-0.001	1109	0.0 %	1109					-1
2.5	2.500	2.453	-0.047	13040	0.0 %	13040					-2
5.0	5.000	5.095	0.095	27020	0.0 %	27020					2
7.5	7.500	7.487	-0.013	39684	0.0 %	39684					0
10.0	10.000	9.973	-0.027	52839	0.0 %	52839					0

LB118552 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Feb 2022 13:46:36

Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	17	-	0.00007471B	PPB	05 Feb 2022 13:48:36	S	Std
0.2 - 1	51.2	1109	-	0.20007471B	PPB	05 Feb 2022 13:50:53	S	Std
2.5 - 1	52.5	13040	-	2.50007471B	PPB	05 Feb 2022 13:53:10	S	Std
5.0 - 1	55	27020	-	5.00007471B	PPB	05 Feb 2022 13:55:27	S	Std
7.5 - 1	57.5	39684	-	7.50007471B	PPB	05 Feb 2022 13:57:43	S	Std
10.0 - 1	510	52839	-	10.00007471B	PPB	05 Feb 2022 14:00:00	S	Std
ICV76 - 1	ICV76	20679	3.8966	-7471B	PPB	05 Feb 2022 14:02:41	U	SMPL
ICB76 - 1	ICB76	-48	-0.0194	-7471B	PPB	05 Feb 2022 14:04:56	U	SMPL
CCV34 - 1	CCV34	26832	5.0591	-7471B	PPB	05 Feb 2022 14:07:12	U	SMPL
CCB34 - 1	CCB34	-60	-0.0217	-7471B	PPB	05 Feb 2022 14:09:26	U	SMPL
CRA - 1	CRA	1092	0.1960	-7471B	PPB	05 Feb 2022 14:11:42	U	SMPL
HighStd - 1	HighStd	52956	9.9948	-7471B	PPB	05 Feb 2022 14:13:57	U	SMPL
ChkStd - 1	ChkStd	36161	6.8217	-7471B	PPB	05 Feb 2022 14:16:13	U	SMPL
PB142509BL - 1	PBS	-71	-0.0238	-7471B	PPB	05 Feb 2022 14:18:30	U	SMPL
PB142509BS - 1	LCSS	19694	3.7105	-7471B	PPB	05 Feb 2022 14:20:48	U	SMPL
N1397-03 - 1	KTS1-MH-02-0-15	18198	3.4279	-7471B	PPB	05 Feb 2022 14:23:04	U	SMPL
N1397-04 - 1	KTS1-MH-01-0-15	75918	14.3331	-7471B	PPB	05 Feb 2022 14:25:21	U	SMPL
N1401-01 - 1	RO-124035	2048	0.3766	-7471B	PPB	05 Feb 2022 14:27:38	U	SMPL
N1401-03 - 1	PSC-124055	2566	0.4745	-7471B	PPB	05 Feb 2022 14:29:56	U	SMPL
N1401-05 - 1	PSC-124042	23983	4.5208	-7471B	PPB	05 Feb 2022 14:32:11	U	SMPL
CCV35 - 1	CCV35	27017	5.0941	-7471B	PPB	05 Feb 2022 14:34:27	U	SMPL
CCB35 - 1	CCB35	-71	-0.0238	-7471B	PPB	05 Feb 2022 14:36:44	U	SMPL
N1401-07 - 1	RO-124030	3002	0.5568	-7471B	PPB	05 Feb 2022 14:39:00	U	SMPL
N1401-07DUP - 1	RO-124030DUP	2757	0.5105	-7471B	PPB	05 Feb 2022 14:41:15	U	SMPL
N1401-07MS - 1	RO-124030MS	24835	4.6818	-7471B	PPB	05 Feb 2022 14:43:30	U	SMPL
N1401-07MSD - 1	RO-124030MSD	26926	5.0769	-7471B	PPB	05 Feb 2022 14:45:46	U	SMPL
N1415-01 - 1	CL-02-020422	2637	0.4879	-7471B	PPB	05 Feb 2022 14:48:02	U	SMPL
N1419-01 - 1	GRIT-STOCKPILE	46078	8.6953	-7471B	PPB	05 Feb 2022 14:50:19	U	SMPL
N1401-07LX5 - 1		431	0.0711	-7471B	PPB	05 Feb 2022 14:52:36	U	SMPL
N1401-07A - 1		24641	4.6452	-7471B	PPB	05 Feb 2022 14:54:56	U	SMPL
N1397-04DLX5 - 1		16184	3.0474	-7471B	PPB	05 Feb 2022 14:57:13	U	SMPL
CCV36 - 1	CCV36	27131	5.1156	-7471B	PPB	05 Feb 2022 14:59:30	U	SMPL
CCB36 - 1	CCB36	-43	-0.0185	-7471B	PPB	05 Feb 2022 15:01:45	U	SMPL