

LB118553


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

40099



A= 0.0000e+000

B= 2.4575e-004

C= -2.6971e-002

Rho= 0.9995237

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	-0.011	-0.011	63	0.000	63					-1
0.2	0.200	0.178	-0.022	834	0.0 %	834					-1
2.5	2.500	2.463	-0.037	10134	0.0 %	10134					1
5.0	5.000	5.030	0.030	20579	0.0 %	20579					3
7.5	7.500	7.712	0.212	31493	0.0 %	31493					-2
10.0	10.000	9.827	-0.173	40099	0.0 %	40099					

LB118553 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 06 Feb 2022 09:54:50				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	63	-	0.0000	7470A	PPB	06 Feb 2022 10:01:03	S Std
0.2 - 1	50.2	834	-	0.2000	7470A	PPB	06 Feb 2022 10:03:20	S Std
2.5 - 1	52.5	10134	-	2.5000	7470A	PPB	06 Feb 2022 10:05:37	S Std
5.0 - 1	55	20579	-	5.0000	7470A	PPB	06 Feb 2022 10:07:54	S Std
7.5 - 1	57.5	31493	-	7.5000	7470A	PPB	06 Feb 2022 10:15:26	S Std
10.0 - 1	518	40099	-	10.0000	7470A	PPB	06 Feb 2022 10:20:23	S Std
ICV77 - 1	ICV77	16522	4.0333	-	7470A	PPB	06 Feb 2022 10:26:08	U SMPL
ICB77 - 1	ICB77	-2	-0.0275	-	7470A	PPB	06 Feb 2022 10:28:23	U SMPL
CCV37 - 1	CCV37	18839	4.6027	-	7470A	PPB	06 Feb 2022 10:30:38	U SMPL
CCB37 - 1	CCB37	11	-0.0243	-	7470A	PPB	06 Feb 2022 10:32:53	U SMPL
CRA - 1	CRA	784	0.1657	-	7470A	PPB	06 Feb 2022 10:37:40	U SMPL
HighStd - 1	HighStd	40149	9.8396	-	7470A	PPB	06 Feb 2022 10:39:55	U SMPL
ChkStd - 1	ChkStd	28063	6.8695	-	7470A	PPB	06 Feb 2022 10:42:10	U SMPL
PB142510BL - 1	PBW	38	-0.0176	-	7470A	PPB	06 Feb 2022 10:44:27	U SMPL
PB142510BS - 1	LCSW	15017	3.6634	-	7470A	PPB	06 Feb 2022 10:46:43	U SMPL
N1399-01 - 1	KTS1-GW-MH-02-8-13	93635	22.9838	-	7470A	PPB	06 Feb 2022 10:48:59	U SMPL
N1400-02 - 1	KTS1-WC-L-MH-02	82271	20.1911	-	7470A	PPB	06 Feb 2022 10:51:16	U SMPL
N1403-01 - 1	MW-1R	2352	0.5510	-	7470A	PPB	06 Feb 2022 10:53:35	U SMPL
N1403-01DUP - 1	MW-1RDUP	2507	0.5891	-	7470A	PPB	06 Feb 2022 10:55:55	U SMPL
N1403-01MS - 1	MW-1RMS	18936	4.6265	-	7470A	PPB	06 Feb 2022 10:58:11	U SMPL
CCV38 - 1	CCV38	19731	4.8219	-	7470A	PPB	06 Feb 2022 11:00:26	U SMPL
CCB38 - 1	CCB38	-9	-0.0292	-	7470A	PPB	06 Feb 2022 11:02:41	U SMPL
N1403-01MSD - 1	MW-1RMSD	18434	4.5032	-	7470A	PPB	06 Feb 2022 11:04:56	U SMPL
N1403-02 - 1	MW-2	344	0.0576	-	7470A	PPB	06 Feb 2022 11:07:11	U SMPL
N1403-03 - 1	MW-3R	3866	0.9231	-	7470A	PPB	06 Feb 2022 11:09:27	U SMPL
N1403-04 - 1	MW-4	159	0.0121	-	7470A	PPB	06 Feb 2022 11:11:42	U SMPL
N1403-05 - 1	MW-4CD	2908	0.6877	-	7470A	PPB	06 Feb 2022 11:13:58	U SMPL
N1403-06 - 1	MW-5	26640	6.5198	-	7470A	PPB	06 Feb 2022 11:16:14	U SMPL
N1403-07 - 1	MW-6	217	0.0264	-	7470A	PPB	06 Feb 2022 11:18:31	U SMPL
N1403-08 - 1	MW-7	245	0.0332	-	7470A	PPB	06 Feb 2022 11:20:49	U SMPL
N1403-09 - 1	MW-12R	132	0.0055	-	7470A	PPB	06 Feb 2022 11:23:06	U SMPL
N1403-10 - 1	MW-13	117	0.0018	-	7470A	PPB	06 Feb 2022 11:25:21	U SMPL
CCV39 - 1	CCV39	19762	4.8295	-	7470A	PPB	06 Feb 2022 11:27:37	U SMPL
CCB39 - 1	CCB39	2	-0.0265	-	7470A	PPB	06 Feb 2022 11:29:52	U SMPL
N1403-11 - 1	MW-14	311	0.0495	-	7470A	PPB	06 Feb 2022 11:32:08	U SMPL
N1403-12 - 1	MW-15	235	0.0308	-	7470A	PPB	06 Feb 2022 11:34:23	U SMPL
N1403-13 - 1	MW-16	76	-0.0083	-	7470A	PPB	06 Feb 2022 11:36:39	U SMPL
N1403-14 - 1	MW-12-A-C-5-A	7199	1.7422	-	7470A	PPB	06 Feb 2022 11:38:54	U SMPL
N1403-01LX5 - 1		349	0.0588	-	7470A	PPB	06 Feb 2022 11:41:10	U SMPL
N1403-01A - 1		18986	4.6388	-	7470A	PPB	06 Feb 2022 11:43:26	U SMPL
N1399-01DLX5 - 1		19588	4.7868	-	7470A	PPB	06 Feb 2022 11:45:42	U SMPL
N1400-02DLX5 - 1		18160	4.4358	-	7470A	PPB	06 Feb 2022 11:48:00	U SMPL
CCV40 - 1	CCV40	19529	4.7723	-	7470A	PPB	06 Feb 2022 11:52:50	U SMPL
CCB40 - 1	CCB40	0	-0.0270	-	7470A	PPB	06 Feb 2022 11:55:05	U SMPL