

LB118554


245.1

INSTRUMENT ID: CV4

Linear

μ Abs.:

36250



A= 0.0000e+000

B= 2.7138e-004

C= -2.3417e-002

Rho= 0.9994913

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.011	-0.011	45	0.000	45					-
0.05	0.050	0.036	-0.014	219	0.0 %	219	0				-28
0.2	0.200	0.182	-0.018	756	0.0 %	756					-9
2.5	2.500	2.403	-0.097	8941	0.0 %	8941					-4
5.0	5.000	5.176	0.176	19159	0.0 %	19159					4
7.5	7.500	7.650	0.150	28276	0.0 %	28276					2
10.0	10.000	9.814	-0.186	36250	0.0 %	36250					-2

LB118554 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 06 Feb 2022 13:43:51

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	45	-	0.0000245.1	PPB	06 Feb 2022 13:45:10	S	Std
0.05 - 1	50.05	219	-	0.0500245.1	PPB	06 Feb 2022 13:47:27	S	Std
0.2 - 1	50.2	756	-	0.2000245.1	PPB	06 Feb 2022 13:49:44	S	Std
2.5 - 1	52.5	8941	-	2.5000245.1	PPB	06 Feb 2022 13:52:01	S	Std
5.0 - 1	55	19159	-	5.0000245.1	PPB	06 Feb 2022 13:54:17	S	Std
7.5 - 1	57.5	28276	-	7.5000245.1	PPB	06 Feb 2022 13:56:33	S	Std
10.0 - 1	510	36250	-	10.0000245.1	PPB	06 Feb 2022 13:58:49	S	Std
ICV78 - 1	ICV78	14999	4.0471	-245.1	PPB	06 Feb 2022 14:04:27	U	SMPL
ICB78 - 1	ICB78	-36	-0.0332	-245.1	PPB	06 Feb 2022 14:06:42	U	SMPL
CCV41 - 1	CCV41	18578	5.0183	-245.1	PPB	06 Feb 2022 14:08:57	U	SMPL
CCB41 - 1	CCB41	-28	-0.0310	-245.1	PPB	06 Feb 2022 14:11:12	U	SMPL
CRA - 1	CRA	229	0.0387	-245.1	PPB	06 Feb 2022 14:21:12	U	SMPL
HighStd - 1	HighStd	36658	9.9249	-245.1	PPB	06 Feb 2022 14:23:27	U	SMPL
ChkStd - 1	ChkStd	25923	7.0116	-245.1	PPB	06 Feb 2022 14:25:42	U	SMPL
PB142511BL - 1	PBW	-29	-0.0313	-245.1	PPB	06 Feb 2022 14:27:59	U	SMPL
PB142511BS - 1	LCSW	13069	3.5233	-245.1	PPB	06 Feb 2022 14:30:15	U	SMPL
N1417-01 - 1	PIT-11-DEWATERING	534	0.1215	-245.1	PPB	06 Feb 2022 14:32:32	U	SMPL
N1417-01DUP - 1	PIT-11-DEWATERINGDUP	554	0.1269	-245.1	PPB	06 Feb 2022 14:34:49	U	SMPL
N1417-01MS - 1	PIT-11-DEWATERINGMS	17246	4.6569	-245.1	PPB	06 Feb 2022 14:37:06	U	SMPL
N1417-01MSD - 1	PIT-11-DEWATERINGMSD	16441	4.4384	-245.1	PPB	06 Feb 2022 14:39:24	U	SMPL
CCV42 - 1	CCV42	19435	5.2509	-245.1	PPB	06 Feb 2022 14:41:39	U	SMPL
CCB42 - 1	CCB42	-30	-0.0316	-245.1	PPB	06 Feb 2022 14:43:55	U	SMPL
N1417-01LX5 - 1		121	0.0094	-245.1	PPB	06 Feb 2022 14:46:10	U	SMPL
N1417-01A - 1		16834	4.5450	-245.1	PPB	06 Feb 2022 14:48:25	U	SMPL
CCV43 - 1	CCV43	19147	5.1728	-245.1	PPB	06 Feb 2022 14:50:40	U	SMPL
CCB43 - 1	CCB43	-37	-0.0335	-245.1	PPB	06 Feb 2022 14:52:55	U	SMPL