



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: Non-Polar Material
Run Number: LB118558
Analysis Date: 02/07/2022
BalanceID: WC SC-4
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 02/07/2022
Extraction IN Time: 08:35
Extraction OUT Time: 09:15
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB118558BL	LB118558BL	WATER	1.3	1000	100	2.9317	2.9317	3.02	2.9318	2.9318	0.0001	0.1
2	LB118558BS	LB118558BS	WATER	1.3	1000	100	3.0179	3.0179	3.01	3.0351	3.0351	0.0172	17.2
3	LB118558BSD	LB118558BSD	WATER	1.3	1000	100	2.9671	2.9671	3.03	2.9844	2.9844	0.0173	17.3
4	N1417-01	PIT-11-DEWATERING	WATER	1.6	1000	100	2.8901	2.8901	3.05	2.8914	2.8914	0.0013	1.3



QC Batch# LB118558 Test: Non-Polar Material Analysis Date: 02/07/2022

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2885
pH Paper 0-14	M4909
Sodium Sulfate	EP2222
1:1 HCL	WP92935
Silica Gel	W2839
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP92936
LCSWD	5.00 ML	WP92937
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	(0.0018-0.0022)	In OVEN TEMP1 :	<u>70 °C</u>	Dessicator Time In1	<u>10:16</u>
1.0000 gram Balance:	<u>1.0005</u>	(0.9950-1.0050)	In Time1:	<u>09:30</u>		
Bal Check Time:	<u>08:55</u>		Out OVEN TEMP1:	<u>70 °C</u>	Dessicator Time Out1	<u>11:00</u>
			Out Time1	<u>10:15</u>		

After Analysis

0.0020 gram Balance:	<u>0.0019</u>	(0.0018-0.0022)	In OVEN TEMP2 :	<u>71 °C</u>	Dessicator Time In2	<u>12:01</u>
1.0000 gram Balance:	<u>1.0004</u>	(0.9950-1.0050)	In Time2:	<u>11:30</u>		
Bal Check Time:	<u>13:00</u>		Out OVEN TEMP2:	<u>71 °C</u>	Dessicator Time Out2	<u>12:30</u>
			Out Time2:	<u>12:00</u>		