



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 2/8/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:15
In Date: 02/07/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 02/08/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB118560

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1413-01	YAWK2	1	1.13	8.44	9.57	8.22	84.0	
N1413-02	YAWK3	2	1.18	8.63	9.81	8.16	80.9	
N1413-03	YAWK4	3	1.15	8.60	9.75	8.29	83.0	
N1413-04	YAWK5	4	1.11	8.49	9.6	8.09	82.2	
N1413-05	YAWK6	5	1.14	8.83	9.97	9.09	90.0	
N1413-06	YAWK7	6	1.18	8.48	9.66	8.3	84.0	
N1413-07	YAWK8	7	1.11	8.78	9.89	8.59	85.2	
N1413-08	YAWK9	8	1.18	8.52	9.7	8.43	85.1	
N1413-09	YAWL0	9	1.15	8.53	9.68	8.56	86.9	
N1413-10	YAWL1	10	1.19	8.49	9.68	8.61	87.4	
N1413-11	YAWL2	11	1.17	8.57	9.74	8.34	83.7	
N1413-12	YAWL3	12	1.15	8.51	9.66	8.1	81.7	
N1413-13	YAWL4	13	1.1	8.79	9.89	8.74	86.9	
N1413-14	YAWL5	14	1.13	8.48	9.61	7.77	78.3	
N1413-15	YAWL6	15	1.18	8.44	9.62	8.13	82.3	
N1413-16	YAWL7	16	1.19	8.52	9.71	8.25	82.9	
N1413-17	YAWL8	17	1.1	8.61	9.71	8.74	88.7	
N1413-18	YAWL9	18	1.15	8.41	9.56	8.16	83.4	
N1413-19	YAWM0	19	1.15	8.81	9.96	8.85	87.4	
N1413-20	YAWMOD	20	1.15	8.81	9.96	8.85	87.4	
N1413-21	YAWMOS	21	1.15	8.81	9.96	8.85	87.4	
N1413-22	YAWM1	22	1.18	8.79	9.97	8.97	88.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$