



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118561

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	02/07/2022	09:25
2	N1366-01		1	12	100.00	>212.0	>212.0	02/07/2022	09:45
3	N1366-01DUP		1	12	100.00	>212.0	>212.0	02/07/2022	10:10
4	N1368-01		1	15	100.00	>212.0	>212.0	02/07/2022	10:40
5	N1392-01		1	10	100.00	>212.0	>212.0	02/07/2022	11:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 118561

WorkList Name : fp2422 WorkList ID : 156582 Department : Wet-Chemistry Date : 02-04-2022 10:43:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/07/2022	Water	N1366-01	C Flash Point	Cool 4 deg C	PSEG03	M11	N28660	02/02/2022	1010B
02/09/2022	Water	N1368-01	F Flash Point	Cool 4 deg C	PSEG03	M11	31382	02/02/2022	1010B
02/10/2022	Water	N1392-01	C Flash Point	Cool 4 deg C	PSEG03	M12	OILY-WATER-NWB-1735	02/03/2022	1010B

Date/Time 02/07/22 09:05
 Raw Sample Received by: KS(wc)
 Raw Sample Relinquished by: KS(wc)

Date/Time 02/07/22 11:20
 Raw Sample Received by: KS(wc)
 Raw Sample Relinquished by: KS(wc)