



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118571

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/07/2022	17:25
2	CAL2	1	Water	NA	NA	20.2	7.00	02/07/2022	17:26
3	CAL3	1	Water	NA	NA	20.2	10.03	02/07/2022	17:27
4	ICV	1	Water	NA	NA	20.4	6.99	02/07/2022	17:29
5	CCV1	1	Water	NA	NA	20.4	2.01	02/07/2022	17:30
6	N1400-01	1	Solid	20.03	20	20.9	6.73	02/07/2022	17:31
7	N1422-02	1	Solid	20.04	20	20.4	10.87	02/07/2022	17:33
8	N1427-02	1	Solid	20.01	20	20.3	9.20	02/07/2022	17:34
9	N1432-02	1	Solid	20.03	20	20.3	5.52	02/07/2022	17:35
10	N1432-05	1	Solid	20.04	20	20.9	6.12	02/07/2022	17:37
11	N1432-08	1	Solid	20.03	20	20.7	5.96	02/07/2022	17:40
12	N1432-11	1	Solid	20.02	20	20.7	6.31	02/07/2022	17:42
13	N1432-14	1	Solid	20.03	20	20.1	6.35	02/07/2022	17:43
14	N1432-17	1	Solid	20.04	20	20.1	6.42	02/07/2022	17:45
15	N1432-20	1	Solid	20.03	20	20.3	6.35	02/07/2022	17:46
16	CCV2	1	Water	NA	NA	20.4	12.03	02/07/2022	17:49
17	N1432-20DUP	1	Solid	20.04	20	20.4	6.36	02/07/2022	17:50
18	CCV3	1	Water	NA	NA	20.4	2.01	02/07/2022	17:51