



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118572

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/07/2022	17:25
2	CAL2	1	Water	NA	NA	20.2	7.00	02/07/2022	17:26
3	CAL3	1	Water	NA	NA	20.2	10.03	02/07/2022	17:27
4	ICV	1	Water	NA	NA	20.4	6.99	02/07/2022	17:29
5	CCV1	1	Water	NA	NA	20.4	2.01	02/07/2022	17:30
6	N1425-03	1	Water	NA	NA	20.3	6.04	02/07/2022	18:00
7	N1425-04	1	Water	NA	NA	20.2	6.56	02/07/2022	18:01
8	N1425-05	1	Water	NA	NA	20.2	6.23	02/07/2022	18:03
9	N1425-07	1	Water	NA	NA	20.3	6.98	02/07/2022	18:05
10	N1425-08	1	Water	NA	NA	20.2	6.04	02/07/2022	18:06
11	N1425-09	1	Water	NA	NA	20.1	6.67	02/07/2022	18:09
12	N1425-10	1	Water	NA	NA	20.3	6.17	02/07/2022	18:10
13	N1425-10DUP	1	Water	NA	NA	20.4	6.18	02/07/2022	18:11
14	CCV2	1	Water	NA	NA	20.8	12.03	02/07/2022	18:12