



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118577

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	02/08/2022	12:45
2	CAL2	1	Water	NA	NA	20.7	7.00	02/08/2022	12:46
3	CAL3	1	Water	NA	NA	20.7	10.03	02/08/2022	12:47
4	ICV	1	Water	NA	NA	20.9	6.99	02/08/2022	12:49
5	CCV1	1	Water	NA	NA	20.7	2.01	02/08/2022	12:50
6	N1434-03	1	Water	NA	NA	20.4	6.13	02/08/2022	12:52
7	N1434-03DUP	1	Water	NA	NA	20.5	6.14	02/08/2022	12:53
8	N1434-04	1	Water	NA	NA	20.3	6.17	02/08/2022	12:55
9	CCV2	1	Water	NA	NA	20.4	12.03	02/08/2022	12:56

W 118577

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1434 W WorkList ID : 156666 Department : Wet-Chemistry Date : 02-08-2022 12:36:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/08/2022	Water	N1434-03	pH	Cool 4 deg C	PSEG03	M11	34040	02/08/2022	9040C
02/08/2022	Water	N1434-04	pH	Cool 4 deg C	PSEG03	M11	34041	02/08/2022	9040C

Date/Time 02/08/22 12:40
 Raw Sample Received by: J P (SM)
 Raw Sample Relinquished by: J P (SM)

Date/Time 02/08/22 13:40
 Raw Sample Received by: J P (SM)
 Raw Sample Relinquished by: J P (SM)