



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118584

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID:

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.9	02/07/2022	17:05
2	N1425-02		1	10	100.00	>212.0	>212.0	02/07/2022	17:35
3	N1425-02DUP		1	10	100.00	>212.0	>212.0	02/07/2022	17:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB118584

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp2722

WorkList ID : 156645

Department : Wet-Chemistry

Date : 02-07-2022 16:45:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/14/2022	Solid	N1425-02	Flash Point	Cool 4 deg C	PSEG03	M11	20622-OIL	02/07/2022	1010A

Date/Time 02/07/22 16:50
Raw Sample Received by: KS(wc)
Raw Sample Relinquished by: KS(wc)

Date/Time 02/07/22 17:55
Raw Sample Received by: KS(wc)
Raw Sample Relinquished by: KS(wc)