



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 2/10/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:00
In Date: 02/09/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:13
Out Date: 02/10/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB118592

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1414-01	YAWM2	1	1.13	8.46	9.59	7.68	77.4	
N1414-02	YAWM3	2	1.13	8.67	9.8	7.74	76.2	
N1414-03	YAWM4	3	1.15	8.82	9.97	8.27	80.7	
N1414-04	YAWM5	4	1.19	8.41	9.6	8.54	87.4	
N1414-05	YAWZ4	5	1.16	8.64	9.8	8.88	89.4	
N1414-06	YAWZ5	6	1.15	8.60	9.75	8.73	88.1	
N1414-07	YAWZ6	7	1.12	8.54	9.66	8.56	87.1	
N1414-08	YAWZ7	8	1.17	8.63	9.8	8.65	86.7	
N1414-09	YAWZ8	9	1.13	8.48	9.61	8.6	88.1	
N1414-10	YAWZ9	10	1.19	8.32	9.51	8.4	86.7	
N1414-11	YAX00	11	1.15	8.76	9.91	8.47	83.6	
N1414-12	YAX01	12	1.14	8.82	9.96	8.42	82.5	
N1414-13	YAX24	13	1.17	8.72	9.89	8.21	80.7	
N1414-14	YAX25	14	1.15	8.79	9.94	8.24	80.7	
N1414-15	YAX26	15	1.16	8.81	9.97	8.25	80.5	
N1414-16	YAX27	16	1.16	8.44	9.6	7.81	78.8	
N1414-17	YAX28	17	1.18	8.57	9.75	7.82	77.5	
N1414-18	YAX29	18	1.1	8.67	9.77	7.81	77.4	
N1414-19	YAX30	19	1.19	8.51	9.7	7.69	76.4	
N1414-20	YAX31	20	1.16	8.58	9.74	7.75	76.8	
N1414-21	YAX31D	21	1.16	8.58	9.74	7.75	76.8	
N1414-22	YAX31S	22	1.16	8.58	9.74	7.75	76.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$