



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118593

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	02/09/2022	09:35
2	N1400-02		1	12	100.00	>212.0	>212.0	02/09/2022	09:55
3	N1400-02DUP		1	12	100.00	>212.0	>212.0	02/09/2022	10:20
4	N1405-05		1	15	100.00	>212.0	>212.0	02/09/2022	10:45
5	N1405-07		1	12	100.00	>212.0	>212.0	02/09/2022	11:05
6	N1417-01		1	13	100.00	>212.0	>212.0	02/09/2022	11:35
7	N1425-06		1	14	100.00	>212.0	>212.0	02/09/2022	12:05
8	N1425-07		1	12	100.00	>212.0	>212.0	02/09/2022	12:35
9	N1425-08		1	15	100.00	>212.0	>212.0	02/09/2022	13:40
10	N1425-09		1	13	100.00	>212.0	>212.0	02/09/2022	14:05
11	N1425-10		1	12	100.00	>212.0	>212.0	02/09/2022	14:40
12	N1431-01		1	15	100.00	>212.0	>212.0	02/09/2022	15:10
13	N1431-02		1	14	100.00	>212.0	>212.0	02/09/2022	15:35
14	N1434-02		1	15	100.00	>212.0	>212.0	02/09/2022	15:58

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp2822 WorkList ID : 156670 Department : Wet-Chemistry Date : 02-08-2022 16:51:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/14/2022	Water	N1425-06	Flash Point	Cool 4 deg C	PSEG03	M11	31377-20422-20722-COMP	02/07/2022	1010B
02/14/2022	Water	N1425-07	Flash Point	Cool 4 deg C	PSEG03	M11	31380	02/07/2022	1010B
02/14/2022	Water	N1425-08	Flash Point	Cool 4 deg C	PSEG03	M11	20522	02/07/2022	1010B
02/10/2022	Water	N1425-09	Flash Point	Cool 4 deg C	PSEG03	M11	20905	02/07/2022	1010B
02/14/2022	Water	N1425-10	Flash Point	Cool 4 deg C	PSEG03	M11	20622	02/07/2022	1010B
02/14/2022	Water	N1431-01	Flash Point	Cool 4 deg C	DVIR01	M11	SHB-2022	02/04/2022	1010B
02/14/2022	Water	N1431-02	Flash Point	Cool 4 deg C	DVIR01	M11	LPBH-2022	02/04/2022	1010B
02/15/2022	Water	N1434-02	Flash Point	Cool 4 deg C	PSEG03	M11	DRUM-COMP	02/08/2022	1010B

H1400-02-65
 H1405-05-F
 H1405-07-F
 H1417-01-D

Date/Time 02/09/22 09:10
 Raw Sample Received by: KSC(COC)
 Raw Sample Relinquished by: RSC(EA-Lab)

Date/Time 02/09/22 16:05
 Raw Sample Received by: RSC(EA-Lab)
 Raw Sample Relinquished by: KSC(WC)